

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE082415\
 Data File : BE090608.D
 Acq On : 24 Aug 2015 19:23
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleID :
 SSTDCCC001EC

Quant Time: Aug 24 23:01:04 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 14 16:52:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	95	0.00
2	1,4-Dioxane	1.000	1.290	-29.0#	104	0.00
3	n-Nitrosodimethylamine	1.000	1.046	-4.6	100	0.00
4 S	2-Fluorophenol	1.000	0.753	24.7	71	0.00
5 S	Phenol-d6	1.000	0.676	32.4#	67	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	87	0.00
7 S	Nitrobenzene-d5	1.000	0.902	9.8	82	0.00
8	Nitrobenzene	1.000	0.873	12.7	81	0.00
9	Naphthalene	1.000	0.977	2.3	85	0.00
10	Hexachlorobutadiene	1.000	1.031	-3.1	89	0.00
11	2-Methylnaphthalene	1.000	0.957	4.3	85	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	95	0.00
13 S	2,4,6-Tribromophenol	1.000	0.842	15.8	88	0.00
14 S	2-Fluorobiphenyl	1.000	0.938	6.2	89	0.00
15	Acenaphthylene	1.000	0.935	6.5	91	0.00
16	Acenaphthene	1.000	0.963	3.7	93	0.00
17	Fluorene	1.000	1.022	-2.2	98	0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	110	0.00
19	4-Bromophenyl-phenylether	1.000	0.926	7.4	105	0.00
20	Hexachlorobenzene	1.000	0.969	3.1	107	0.00
21	Pentachlorophenol	1.000	0.813	18.7	109	0.00
22	Phenanthrene	1.000	0.950	5.0	106	0.00
23	Anthracene	1.000	0.932	6.8	106	0.00
24	Fluoranthene	1.000	0.987	1.3	111	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	125	0.00
26	Pyrene	1.000	0.906	9.4	116	0.00
27 S	Terphenyl-d14	1.000	0.904	9.6	115	0.00
28	Benzo(a)anthracene	1.000	0.920	8.0	121	0.00
29	Chrysene	1.000	0.998	0.2	126	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.649	35.1#	86	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.888	11.2	120	0.01
32 I	Perylene-d12	5.000	5.000	0.0	122	0.00
33	Benzo(b)fluoranthene	1.000	0.939	6.1	117	0.00
34	Benzo(k)fluoranthene	1.000	1.035	-3.5	125	0.00
35 C	Benzo(a)pyrene	1.000	0.943	5.7	118	0.00
36	Dibenzo(a,h)anthracene	1.000	0.912	8.8	117	0.00
37	Benzo(g,h,i)perylene	1.000	0.927	7.3	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0