

Data Path : Z:\HPCHEM1\BNA_E\Data\BE082515\
 Data File : BE090610.D
 Acq On : 25 Aug 2015 12:23
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 26 03:25:40 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 14 16:52:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.658	0.799	-21.4	120	0.00
3	n-Nitrosodimethylamine	0.863	0.882	-2.2	108	0.00
4 S	2-Fluorophenol	1.446	1.104	23.7	79	0.00
5 S	Phenol-d6	2.125	1.539	27.6#	79	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
7 S	Nitrobenzene-d5	0.398	0.370	7.0	94	0.00
8	Nitrobenzene	0.413	0.371	10.2	92	0.00
9	Naphthalene	1.136	1.115	1.8	95	0.00
10	Hexachlorobutadiene	0.167	0.171	-2.4	98	0.00
11	2-Methylnaphthalene	0.740	0.703	5.0	93	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
13 S	2,4,6-Tribromophenol	0.192	0.160	16.7	95	0.00
14 S	2-Fluorobiphenyl	1.493	1.406	5.8	98	0.00
15	Acenaphthylene	9.132	8.587	6.0	99	0.00
16	Acenaphthene	1.374	1.325	3.6	101	0.00
17	Fluorene	1.684	1.769	-5.0	110	0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
19	4-Bromophenyl-phenylether	0.204	0.181	11.3	109	0.00
20	Hexachlorobenzene	0.906	0.885	2.3	118	-0.02
21	Pentachlorophenol	0.079	0.058	26.6#	117	0.00
22	Phenanthrene	1.312	1.247	5.0	116	0.00
23	Anthracene	1.208	1.144	5.3	118	0.00
24	Fluoranthene	1.396	1.422	-1.9	126	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	141	0.00
26	Pyrene	1.430	1.287	10.0	130	0.00
27 S	Terphenyl-d14	0.914	0.818	10.5	128	0.00
28	Benzo(a)anthracene	1.316	1.235	6.2	139	0.00
29	Chrysene	1.344	1.336	0.6	142	0.00
30	Bis(2-ethylhexyl)phthalate	0.725	0.450	37.9#	107	0.00
31	Indeno(1,2,3-cd)pyrene	1.266	1.132	10.6	137	0.00
32 I	Perylene-d12	1.000	1.000	0.0	140	0.00
33	Benzo(b)fluoranthene	1.184	1.170	1.2	142	0.00
34	Benzo(k)fluoranthene	1.342	1.335	0.5	138	0.00
35 C	Benzo(a)pyrene	1.108	1.052	5.1	138	0.00
36	Dibenzo(a,h)anthracene	1.069	0.962	10.0	133	0.00
37	Benzo(g,h,i)perylene	1.169	1.085	7.2	133	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0