

Data Path : Z:\HPCHEM1\BNA_E\Data\BE082515\
 Data File : BE090625.D
 Acq On : 25 Aug 2015 22:11
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 26 03:28:34 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 14 16:52:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	101	0.00
2	1,4-Dioxane	1.000	1.159	-15.9	100	0.00
3	n-Nitrosodimethylamine	1.000	1.034	-3.4	106	0.00
4 S	2-Fluorophenol	1.000	0.760	24.0	76	0.00
5 S	Phenol-d6	1.000	0.692	30.8#	73	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	93	0.00
7 S	Nitrobenzene-d5	1.000	0.934	6.6	91	0.00
8	Nitrobenzene	1.000	0.897	10.3	89	0.00
9	Naphthalene	1.000	0.977	2.3	91	0.00
10	Hexachlorobutadiene	1.000	1.040	-4.0	97	0.00
11	2-Methylnaphthalene	1.000	0.949	5.1	90	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	102	0.00
13 S	2,4,6-Tribromophenol	1.000	0.879	12.1	99	0.00
14 S	2-Fluorobiphenyl	1.000	0.932	6.8	95	0.00
15	Acenaphthylene	1.000	0.940	6.0	98	0.00
16	Acenaphthene	1.000	0.989	1.1	102	0.00
17	Fluorene	1.000	1.052	-5.2	108	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	118	-0.02
19	4-Bromophenyl-phenylether	1.000	0.895	10.5	109	0.00
20	Hexachlorobenzene	1.000	0.995	0.5	118	-0.02
21	Pentachlorophenol	1.000	0.862	13.8	128	0.00
22	Phenanthrene	1.000	0.949	5.1	114	0.00
23	Anthracene	1.000	0.950	5.0	116	0.00
24	Fluoranthene	1.000	0.998	0.2	121	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	136	0.00
26	Pyrene	1.000	0.908	9.2	126	0.00
27 S	Terphenyl-d14	1.000	0.904	9.6	125	0.00
28	Benzo(a)anthracene	1.000	0.916	8.4	131	0.00
29	Chrysene	1.000	1.021	-2.1	140	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.704	29.6#	104	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.896	10.4	132	0.00
32 I	Perylene-d12	5.000	5.000	0.0	135	0.00
33	Benzo(b)fluoranthene	1.000	0.933	6.7	129	0.00
34	Benzo(k)fluoranthene	1.000	0.994	0.6	133	0.00
35 C	Benzo(a)pyrene	1.000	0.940	6.0	131	0.00
36	Dibenzo(a,h)anthracene	1.000	0.898	10.2	128	0.00
37	Benzo(g,h,i)perylene	1.000	0.917	8.3	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0