

Data Path : Z:\HPCHEM1\BNA_E\Data\BE082515\
 Data File : BE090627.D
 Acq On : 25 Aug 2015 23:23
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 26 03:28:52 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 14 16:52:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.658	0.704	-7.0	103	0.00
3	n-Nitrosodimethylamine	0.863	0.877	-1.6	105	0.00
4 S	2-Fluorophenol	1.446	1.075	25.7#	75	0.00
5 S	Phenol-d6	2.125	1.440	32.2#	72	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
7 S	Nitrobenzene-d5	0.398	0.351	11.8	87	0.00
8	Nitrobenzene	0.413	0.365	11.6	89	0.00
9	Naphthalene	1.136	1.111	2.2	92	0.00
10	Hexachlorobutadiene	0.167	0.173	-3.6	98	0.00
11	2-Methylnaphthalene	0.740	0.686	7.3	89	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
13 S	2,4,6-Tribromophenol	0.192	0.165	14.1	95	0.00
14 S	2-Fluorobiphenyl	1.493	1.383	7.4	93	0.00
15	Acenaphthylene	9.132	8.561	6.3	96	0.00
16	Acenaphthene	1.374	1.348	1.9	100	0.00
17	Fluorene	1.684	1.731	-2.8	104	0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	115	-0.02
19	4-Bromophenyl-phenylether	0.204	0.185	9.3	107	0.00
20	Hexachlorobenzene	0.906	0.881	2.8	113	-0.02
21	Pentachlorophenol	0.079	0.063	20.3	122	0.00
22	Phenanthrene	1.312	1.245	5.1	112	0.00
23	Anthracene	1.208	1.135	6.0	112	0.00
24	Fluoranthene	1.396	1.379	1.2	117	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	133	0.00
26	Pyrene	1.430	1.278	10.6	121	0.00
27 S	Terphenyl-d14	0.914	0.809	11.5	119	0.00
28	Benzo(a)anthracene	1.316	1.205	8.4	128	0.00
29	Chrysene	1.344	1.387	-3.2	138	0.00
30	Bis(2-ethylhexyl)phthalate	0.725	0.436	39.9#	97	0.00
31	Indeno(1,2,3-cd)pyrene	1.266	1.169	7.7	133	0.00
32 I	Perylene-d12	1.000	1.000	0.0	134	0.00
33	Benzo(b)fluoranthene	1.184	1.113	6.0	129	0.00
34	Benzo(k)fluoranthene	1.342	1.395	-3.9	138	0.00
35 C	Benzo(a)pyrene	1.108	1.048	5.4	131	0.00
36	Dibenzo(a,h)anthracene	1.069	0.973	9.0	129	0.00
37	Benzo(g,h,i)perylene	1.169	1.089	6.8	127	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0