


Data Path : Z:\HPCHEM1\BNA_E\Data\BE082515\
Data File : BE090609.D
Acq On : 25 Aug 2015 11:45
Operator : UM/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE081415.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Aug 14 16:52:42 2015



AutoFind: Scans 2516, 2517, 2518; Background Corrected with Scan 2508

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	55.1	32015	PASS
68	69	0.00	2	1.1	334	PASS
69	198	0.00	100	52.9	30712	PASS
70	69	0.00	2	0.4	132	PASS
127	198	10	80	56.3	32709	PASS
197	198	0.00	2	0.2	91	PASS
198	198	100	100	100.0	58051	PASS
199	198	5	9	6.6	3809	PASS
275	198	10	60	25.5	14830	PASS
365	198	1	100	3.4	1956	PASS
441	442	0.01	24	16.3	8710	PASS
442	198	50	100	91.8	53282	PASS
443	442	15	24	19.3	10285	PASS