

Data Path : Z:\HPCHEM1\BNA E\DATA\BE082615\
 Data File : BE090635.D
 Acq On : 26 Aug 2015 11:52
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 26 23:41:22 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 14 16:52:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.658	0.700	-6.4	109	0.00
3	n-Nitrosodimethylamine	0.863	0.880	-2.0	112	0.00
4 S	2-Fluorophenol	1.446	1.060	26.7#	79	0.00
5 S	Phenol-d6	2.125	1.469	30.9#	78	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
7 S	Nitrobenzene-d5	0.398	0.373	6.3	99	0.00
8	Nitrobenzene	0.413	0.373	9.7	97	0.00
9	Naphthalene	1.136	1.106	2.6	98	0.00
10	Hexachlorobutadiene	0.167	0.173	-3.6	104	0.00
11	2-Methylnaphthalene	0.740	0.687	7.2	95	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
13 S	2,4,6-Tribromophenol	0.192	0.155	19.3	97	0.00
14 S	2-Fluorobiphenyl	1.493	1.348	9.7	99	0.00
15	Acenaphthylene	9.132	8.472	7.2	104	0.00
16	Acenaphthene	1.374	1.342	2.3	108	0.00
17	Fluorene	1.684	1.767	-4.9	116	0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	127	0.00
19	4-Bromophenyl-phenylether	0.204	0.179	12.3	114	0.00
20	Hexachlorobenzene	0.906	0.874	3.5	123	-0.02
21	Pentachlorophenol	0.079	0.056	29.1#	119	0.00
22	Phenanthrene	1.312	1.231	6.2	121	0.00
23	Anthracene	1.208	1.115	7.7	121	0.00
24	Fluoranthene	1.396	1.357	2.8	126	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	143	0.00
26	Pyrene	1.430	1.287	10.0	132	0.00
27 S	Terphenyl-d14	0.914	0.799	12.6	127	0.00
28	Benzo(a)anthracene	1.316	1.203	8.6	138	0.00
29	Chrysene	1.344	1.385	-3.1	149	0.00
30	Bis(2-ethylhexyl)phthalate	0.725	0.423	41.7#	102	0.00
31	Indeno(1,2,3-cd)pyrene	1.266	1.156	8.7	142	0.00
32 I	Perylene-d12	1.000	1.000	0.0	143	0.00
33	Benzo(b)fluoranthene	1.184	1.173	0.9	146	0.00
34	Benzo(k)fluoranthene	1.342	1.354	-0.9	143	0.00
35 C	Benzo(a)pyrene	1.108	1.048	5.4	140	0.00
36	Dibenzo(a,h)anthracene	1.069	0.977	8.6	138	0.00
37	Benzo(g,h,i)perylene	1.169	1.084	7.3	135	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0