

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE082715\
 Data File : BE090647.D
 Acq On : 27 Aug 2015 12:23
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 27 22:51:02 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 14 16:52:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2	1,4-Dioxane	0.658	0.706	-7.3	90	-0.01
3	n-Nitrosodimethylamine	0.863	0.820	5.0	85	0.00
4 S	2-Fluorophenol	1.446	0.990	31.5#	60	0.00
5 S	Phenol-d6	2.125	1.368	35.6#	59	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	81	0.00
7 S	Nitrobenzene-d5	0.398	0.327	17.8	70	0.00
8	Nitrobenzene	0.413	0.335	18.9	70	0.00
9	Naphthalene	1.136	1.064	6.3	76	0.00
10	Hexachlorobutadiene	0.167	0.167	0.0	81	0.00
11	2-Methylnaphthalene	0.740	0.642	13.2	72	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
13 S	2,4,6-Tribromophenol	0.192	0.134	30.2#	68	0.00
14 S	2-Fluorobiphenyl	1.493	1.272	14.8	75	0.00
15	Acenaphthylene	9.132	7.849	14.0	77	0.00
16	Acenaphthene	1.374	1.284	6.6	84	0.00
17	Fluorene	1.684	1.673	0.7	88	0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
19	4-Bromophenyl-phenylether	0.204	0.179	12.3	91	0.00
20	Hexachlorobenzene	0.906	0.866	4.4	97	-0.02
21	Pentachlorophenol	0.079	0.046	41.8#	79	0.00
22	Phenanthrene	1.312	1.208	7.9	95	0.00
23	Anthracene	1.208	1.077	10.8	93	0.00
24	Fluoranthene	1.396	1.268	9.2	94	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
26	Pyrene	1.430	1.217	14.9	99	0.00
27 S	Terphenyl-d14	0.914	0.757	17.2	95	0.00
28	Benzo(a)anthracene	1.316	1.077	18.2	97	0.00
29	Chrysene	1.344	1.333	0.8	113	0.00
30	Bis(2-ethylhexyl)phthalate	0.725	0.322	55.6#	61	0.00
31	Indeno(1,2,3-cd)pyrene	1.266	0.947	25.2#	92	0.02
32 I	Perylene-d12	1.000	1.000	0.0	108	0.00
33	Benzo(b)fluoranthene	1.184	1.033	12.8	97	0.00
34	Benzo(k)fluoranthene	1.342	1.280	4.6	102	0.00
35 C	Benzo(a)pyrene	1.108	0.922	16.8	93	0.00
36	Dibenzo(a,h)anthracene	1.069	0.834	22.0	89	0.01
37	Benzo(g,h,i)perylene	1.169	0.969	17.1	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0