

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE082715\
 Data File : BE090654.D
 Acq On : 27 Aug 2015 17:24
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 27 23:05:07 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 27 23:02:45 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2	1,4-Dioxane	0.658	0.697	-5.9	124	0.00
3	n-Nitrosodimethylamine	0.863	0.816	5.4	119	0.00
4 S	2-Fluorophenol	1.446	0.996	31.1#	85	0.00
5 S	Phenol-d6	2.125	1.332	37.3#	81	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
7 S	Nitrobenzene-d5	0.398	0.336	15.6	100	0.00
8	Nitrobenzene	0.413	0.338	18.2	99	0.00
9	Naphthalene	1.136	1.052	7.4	105	0.00
10	Hexachlorobutadiene	0.167	0.168	-0.6	114	0.00
11	2-Methylnaphthalene	0.740	0.639	13.6	100	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
13 S	2,4,6-Tribromophenol	0.192	0.143	25.5#	101	0.00
14 S	2-Fluorobiphenyl	1.493	1.268	15.1	104	0.00
15	Acenaphthylene	9.132	7.879	13.7	108	0.00
16	Acenaphthene	1.374	1.285	6.5	116	0.00
17	Fluorene	1.684	1.674	0.6	123	0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	141	0.00
19	4-Bromophenyl-phenylether	0.204	0.171	16.2	122	0.00
20	Hexachlorobenzene	0.906	0.848	6.4	133	-0.02
21	Pentachlorophenol	0.079	0.052	34.2#	123	0.00
22	Phenanthrene	1.312	1.194	9.0	131	0.00
23	Anthracene	1.208	1.090	9.8	132	0.00
24	Fluoranthene	1.396	1.258	9.9	131	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	158#	0.00
26	Pyrene	1.430	1.207	15.6	136	0.00
27 S	Terphenyl-d14	0.914	0.760	16.8	133	0.00
28	Benzo(a)anthracene	1.316	1.112	15.5	140	0.00
29	Chrysene	1.344	1.271	5.4	150#	0.00
30	Bis(2-ethylhexyl)phthalate	0.725	0.328	54.8#	87	0.00
31	Indeno(1,2,3-cd)pyrene	1.266	0.931	26.5#	125	0.02
32 I	Perylene-d12	1.000	1.000	0.0	149	0.00
33	Benzo(b)fluoranthene	1.184	1.037	12.4	134	0.00
34	Benzo(k)fluoranthene	1.342	1.269	5.4	140	0.00
35 C	Benzo(a)pyrene	1.108	0.925	16.5	128	0.00
36	Dibenzo(a,h)anthracene	1.069	0.826	22.7	121	0.00
37	Benzo(g,h,i)perylene	1.169	0.957	18.1	124	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0