

Data Path : Z:\HPCHEM1\BNA_E\Data\BE082817\
 Data File : BE093889.D
 Acq On : 28 Aug 2017 9:27
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 28 11:53:21 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 28 11:26:07 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.550	0.540	1.8	107	0.00
3	n-Nitrosodimethylamine	0.612	0.607	0.8	107	0.00
4 S	2-Fluorophenol	1.231	1.178	4.3	100	0.00
5 S	Phenol-d6	1.882	1.771	5.9	101	0.00
6	bis(2-Chloroethyl)ether	1.450	1.424	1.8	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.333	0.315	5.4	100	0.00
9	Naphthalene	4.555	4.472	1.8	102	0.00
10	Hexachlorobutadiene	0.161	0.157	2.5	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.622	1.7	103	0.00
12	2-Methylnaphthalene	0.761	0.737	3.2	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.107	0.092	14.0	100	0.00
15 S	2-Fluorobiphenyl	1.592	1.558	2.1	101	-0.01
16	Acenaphthylene	1.979	1.883	4.9	103	0.00
17	Acenaphthene	1.335	1.303	2.4	103	0.00
18	Fluorene	1.727	1.696	1.8	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
20	4-Bromophenyl-phenylether	0.166	0.147	11.4	101	0.00
21	Hexachlorobenzene	0.144	0.136	5.6	103	-0.03
22	Pentachlorophenol	0.063	0.052	17.5	115	0.00
23	Phenanthrene	1.017	0.926	8.9	107	-0.03
24	Anthracene	1.212	1.064	12.2	108	0.00
25 SURR	Fluoranthene-d10	4.612	4.410	4.4	104	0.00
26	Fluoranthene	1.393	1.330	4.5	105	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
28	Pyrene	1.346	1.303	3.2	105	0.00
29 S	Terphenyl-d14	0.724	0.700	3.3	104	0.00
30	Benzo(a)anthracene	1.188	1.113	6.3	103	-0.01
31	Chrysene	1.320	1.263	4.3	102	0.00
32	Bis(2-ethylhexyl)phthalate	2.504	2.182	12.9	99	0.00
33	Indeno(1,2,3-cd)pyrene	0.901	0.823	8.7	100	0.00
34 I	Perylene-d12	1.000	1.000	0.0	104	0.00
35	Benzo(b)fluoranthene	1.277	1.135	11.1	102	0.00
36	Benzo(k)fluoranthene	1.460	1.345	7.9	98	0.00
37 C	Benzo(a)pyrene	1.268	1.189	6.2	104	0.00
38	Dibenzo(a,h)anthracene	0.895	0.805	10.1	99	0.00
39	Benzo(g,h,i)perylene	0.975	0.900	7.7	103	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE082817\
Data File : BE093889.D
Acq On : 28 Aug 2017 9:27
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 28 11:53:21 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 28 11:26:07 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		