

Data Path : Z:\HPCHEM1\BNA E\DATA\BE082916\
 Data File : BE092293.D
 Acq On : 29 Aug 2016 17:37
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE082916

Quant Time: Aug 29 18:43:42 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE082916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 29 18:19:23 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	101	0.00
2	1,4-Dioxane	0.400	0.445	-11.2	108	0.00
3	n-Nitrosodimethylamine	0.400	0.422	-5.5	122	-0.02
4 S	2-Fluorophenol	0.400	0.396	1.0	119	-0.02
5 S	Phenol-d6	0.400	0.348	13.0	111	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.428	-7.0	128	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	105	-0.02
8 S	Nitrobenzene-d5	0.400	0.333	16.8	117	0.00
9	Naphthalene	0.400	0.393	1.8	103	0.00
10	Hexachlorobutadiene	0.400	0.403	-0.8	105	0.00
11	2-Methylnaphthalene	0.400	0.389	2.8	107	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	106	0.00
13 S	2,4,6-Tribromophenol	0.400	0.360	10.0	106	0.00
14 S	2-Fluorobiphenyl	0.400	0.393	1.8	108	0.00
15	Acenaphthylene	0.400	0.379	5.3	107	0.00
16	Acenaphthene	0.400	0.390	2.5	107	0.00
17	Fluorene	0.400	0.389	2.8	108	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	107	0.00
19	Hexachlorobenzene	0.400	0.378	5.5	106	0.00
20	Pentachlorophenol	0.400	0.429	-7.2	118	0.00
21	Phenanthrene	0.400	0.395	1.3	110	0.00
22	Anthracene	0.400	0.378	5.5	108	0.00
23	Fluoranthene	0.400	0.364	9.0	105	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	107	0.00
25	Pyrene	0.400	0.381	4.8	108	0.00
26 S	Terphenyl-d14	0.400	0.383	4.3	104	0.00
27	Benzo(a)anthracene	0.400	0.455	-13.7	107	0.00
28	Chrysene	0.400	0.418	-4.5	104	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.353	11.8	103	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.383	4.3	111	0.00
31 I	Perylene-d12	5.000	5.000	0.0	104	0.00
32	Benzo(b)fluoranthene	0.400	0.388	3.0	102	0.00
33	Benzo(k)fluoranthene	0.400	0.385	3.8	107	0.00
34 C	Benzo(a)pyrene	0.400	0.374	6.5	105	0.01
35	Dibenzo(a,h)anthracene	0.400	0.379	5.3	113	0.00
36	Benzo(g,h,i)perylene	0.400	0.386	3.5	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0