

Data Path : Z:\HPCHEM1\BNA E\DATA\BE082916\
 Data File : BE092295.D
 Acq On : 29 Aug 2016 18:55
 Operator : UM/SJ
 Sample : PB93078BS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB93078BS

Manual Integrations
 APPROVED

sohil
 8/30/2016 4:48:24 PM

Quant Time: Aug 30 00:55:23 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE082916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 29 18:19:23 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.19	152	15454	5.00	ng	0.00
7) Naphthalene-d8	10.95	136	66090	5.00	ng	-0.02
12) Acenaphthene-d10	14.83	164	31439	5.00	ng	0.00
18) Phenanthrene-d10	17.58	188	64113	5.00	ng	0.00
24) Chrysene-d12	21.75	240	69105	5.00	ng	0.00
31) Perylene-d12	24.12	264	62196	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.68	112	241784	69.73	ng	-0.02
5) Phenol-d6	7.33	99	298491	65.35	ng	-0.05
8) Nitrobenzene-d5	9.34	82	297065	68.70	ng	-0.02
13) 2,4,6-Tribromophenol	16.31	330	62171	73.02	ng	-0.02
14) 2-Fluorobiphenyl	13.45	172	516863	52.36	ng	-0.01
26) Terphenyl-d14	20.19	244	527669	56.51	ng	0.00

Target Compounds						Qvalue
2) 1,4-Dioxane	3.43	88	4356	2.22	ng	# 85
3) n-Nitrosodimethylamine	3.78	42	7848	3.43	ng	# 97
6) bis(2-Chloroethyl)ether	7.59	93	11547	3.25	ng	91
9) Naphthalene	11.01	128	38028	2.58	ng	95
10) Hexachlorobutadiene	11.30	225	5811	2.60	ng	99
11) 2-Methylnaphthalene	12.64	142	25217	2.77	ng	93
15) Acenaphthylene	14.53	152	143712	2.67	ng	99
16) Acenaphthene	14.88	154	21801	2.48	ng	95
17) Fluorene	15.87	166	27183	2.56	ng	99
19) Hexachlorobenzene	16.89	284	7562	2.87	ng	# 100
20) Pentachlorophenol	17.23	266	5475	5.15	ng	# 80
21) Phenanthrene	17.60	178	39774	2.64	ng	# 81
22) Anthracene	17.69	178	37529	2.71	ng	# 83
23) Fluoranthene	19.61	202	178814	2.65	ng	100
25) Pyrene	19.97	202	182178	2.85	ng	99
27) Benzo(a)anthracene	21.72	228	195883m	2.71	ng	
28) Chrysene	21.79	228	162144m	2.50	ng	
29) Bis(2-ethylhexyl)phthalate	21.66	149	12939	2.85	ng	# 95
30) Indeno(1,2,3-cd)pyrene	26.62	276	205042	3.11	ng	99
32) Benzo(b)fluoranthene	23.38	252	41310	2.80	ng	# 95
33) Benzo(k)fluoranthene	23.43	252	44593	2.67	ng	# 94
34) Benzo(a)pyrene	24.01	252	41181	2.79	ng	# 93
35) Dibenzo(a,h)anthracene	26.63	278	42855	3.31	ng	93
36) Benzo(g,h,i)perylene	27.38	276	182534	3.15	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE082916\
Data File : BE092295.D
Acq On : 29 Aug 2016 18:55
Operator : UM/SJ
Sample : PB93078BS
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
PB93078BS

Manual Integrations
APPROVED

sohil
8/30/2016 4:48:24 PM

Quant Time: Aug 30 00:55:23 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE082916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 29 18:19:23 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.19 min Scan# 577

Delta R.T. 0.00 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Instrument :

BNA_E





ClientSampleId :

PB93078BS

Tot Ion	Ratio	Lower	Upper
152	100		
150	140.4	135.6	203.4
115	43.4	59.4	89.0

Manual Integrations
APPROVED

sohil
8/30/2016 4:48:24 PM



#2

1,4-Dioxane

Concen: 2.22 ng




RT: 3.43 min Scan# 73




Delta R.T. 0.00 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Lower	Upper
88	100		
58	90.5	64.4	96.6
43	55.8	33.4	50.2









#3
n-Nitrosodimethylamine
Concen: 3.43 ng
RT: 3.78 min Scan# 103
Delta R.T. -0.03 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tot Ion	Ratio	Lower	Upper
42	100		
74	94.5	77.2	115.8
44	10.4	5.0	7.4

Instrument :
BNA_E
ClientSampleId :
PB93078BS


Manual Integrations
APPROVED




sohil
8/30/2016 4:48:24 PM



#4
2-Fluorophenol
Concen: 69.73 ng
RT: 5.68 min Scan# 365
Delta R.T. -0.02 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Lower	Upper
112	100		
64	69.3	53.2	79.8
63	38.1	27.4	41.2









#5
Phenol-d6
Concen: 65.35 ng
RT: 7.33 min Scan# 540
Delta R.T. -0.05 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tot Ion	Ratio	Resp	Lower	Upper
99	100	298491		
42	28.9	0.0	0.0	0.0
71	38.3	29.8	44.8	

Instrument :
BNA_E
ClientSampleId :
PB93078BS


Manual Integrations
APPROVED


sohil
8/30/2016 4:48:24 PM



#6
bis(2-Chloroethyl)ether
Concen: 3.25 ng
RT: 7.59 min Scan# 551
Delta R.T. -0.02 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	11547		
63	96.5	68.5	102.7	
95	32.3	27.2	40.8	





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.95 min Scan# 708

Delta R.T. -0.02 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Instrument :

BNA_E





ClientSampleId :

PB93078BS

Tot Ion	Ratio	Resp	Lower	Upper
136	100	66090		
137	10.0	8.8	13.2	
54	13.2	6.2	9.4#	
68	8.6	4.6	6.8#	

Manual Integrations
APPROVED

sohil
8/30/2016 4:48:24 PM



#8

Nitrobenzene-d5

Concen: 68.70 ng




RT: 9.34 min Scan# 627


Delta R.T. -0.02 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	297065		
128	21.5	45.0	67.4#	
54	63.2	45.4	68.2	





#9

Naphthalene

Concen: 2.58 ng

RT: 11.01 min Scan# 711

Delta R.T. 0.00 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Instrument :

BNA_E





ClientSampleId :

PB93078BS

Tot Ion:	128	Resp:	38028
Ion Ratio	Lower	Upper	
128	100		
129	11.6	11.5	17.3
127	12.3	11.0	16.6

Manual Integrations
APPROVED

sohil
8/30/2016 4:48:24 PM



#10

Hexachlorobutadiene

Concen: 2.60 ng




RT: 11.30 min Scan# 726


Delta R.T. 0.00 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Tgt Ion:	225	Resp:	5811
Ion Ratio	Lower	Upper	
225	100		
223	63.7	52.2	78.4
227	63.2	51.0	76.6









#11
2-Methylnaphthalene
Concen: 2.77 ng
RT: 12.64 min Scan# 815
Delta R.T. -0.02 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Instrument :
BNA_E
ClientSampleId :
PB93078BS

Tot Ion	Ratio	Lower	Upper
142	100		
141	83.6	70.6	105.8
115	32.1	31.0	46.4




Manual Integrations
APPROVED


sohil
8/30/2016 4:48:24 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.83 min Scan# 981
Delta R.T. 0.00 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Lower	Upper
164	100		
162	91.5	83.3	124.9
160	35.6	38.0	57.0#









#13
 2,4,6-Tribromophenol
 Concen: 73.02 ng
 RT: 16.31 min Scan# 1088
 Delta R.T. -0.02 min
 Lab File: BE092295.D
 Acq: 29 Aug 2016 18:55

Instrument :
 BNA_E
 ClientSampled :
 PB93078BS

Tot Ion:330 Resp: 62171
 Ion Ratio Lower Upper
 330 100
 332 97.5 75.4 113.0
 141 45.2 37.7 56.5




Manual Integrations
 APPROVED


sohil
 8/30/2016 4:48:24 PM



#14
 2-Fluorobiphenyl
 Concen: 52.36 ng
 RT: 13.45 min Scan# 885
 Delta R.T. -0.01 min
 Lab File: BE092295.D
 Acq: 29 Aug 2016 18:55

Tgt Ion:172 Resp: 516863
 Ion Ratio Lower Upper
 172 100
 171 34.9 26.8 40.2
 170 23.0 17.8 26.6





#15

Acenaphthylene

Concen: 2.67 ng

RT: 14.53 min Scan# 963

Delta R.T. -0.02 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Instrument :

BNA_E





ClientSampleId :

PB93078BS

Tot Ion:	152	Resp:	143712
Ion Ratio	Lower	Upper	
152	100		
151	4.9	3.8	5.8
153	13.3	10.2	15.2

Manual Integrations
APPROVED

sohil
8/30/2016 4:48:24 PM



#16

Acenaphthene

Concen: 2.48 ng




RT: 14.88 min Scan# 984


Delta R.T. 0.00 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Tgt Ion:	154	Resp:	21801
Ion Ratio	Lower	Upper	
154	100		
153	458.6	355.0	532.4
152	229.7	179.7	269.5









#17
Fluorene
 Concen: 2.56 ng
 RT: 15.87 min Scan# 1050
 Delta R.T. -0.01 min
 Lab File: BE092295.D
 Acq: 29 Aug 2016 18:55

Instrument :
 BNA_E
 ClientSampleId :
 PB93078BS

Tot Ion	Ratio	Lower	Upper
166	100		
165	98.5	80.1	120.1
167	13.4	10.6	16.0




Manual Integrations
 APPROVED


sohil
 8/30/2016 4:48:24 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.58 min Scan# 1154
 Delta R.T. -0.00 min
 Lab File: BE092295.D
 Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Lower	Upper
188	100		
94	3.9	9.2	13.8#
80	3.2	9.1	13.7#









#19
Hexachlorobenzene
Concen: 2.87 ng
RT: 16.89 min Scan# 1124
Delta R.T. -0.00 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Instrument :
BNA_E
ClientSampleId :
PB93078BS

Tot Ion	Ratio	Resp	Lower	Upper
284	100	7562		
142	0.0	0.0	0.0	0.0
249	33.5	0.0	0.0	0.0




Manual Integrations
APPROVED


sohil
8/30/2016 4:48:24 PM



#20
Pentachlorophenol
Concen: 5.15 ng
RT: 17.23 min Scan# 1139
Delta R.T. -0.02 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	5475		
268	56.4	62.3	93.5	
264	69.9	67.4	101.0	





#21

Phenanthrene

Concen: 2.64 ng

RT: 17.60 min Scan# 1155

Delta R.T. -0.02 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Instrument :

BNA_E





ClientSampleId :

PB93078BS

Tot Ion	Ratio	Lower	Upper
178	100		
176	20.2	15.1	22.7
179	0.0	12.9	19.3#

Manual Integrations
APPROVED

sohil
8/30/2016 4:48:24 PM



#22

Anthracene

Concen: 2.71 ng




RT: 17.69 min Scan# 1159


Delta R.T. -0.02 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.0	22.4
179	0.0	12.0	18.0#





#23

Fluoranthene

Concen: 2.65 ng

RT: 19.61 min Scan# 1266

Delta R.T. -0.02 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Instrument :

BNA_E





ClientSampleId :

PB93078BS

Tot Ion:	202	Resp:	178814
Ion Ratio	Lower	Upper	
202	100		
101	2.7	2.2	3.4
203	17.2	13.7	20.5

Manual Integrations
APPROVED

sohil
8/30/2016 4:48:24 PM



#24

Chrysene-d12

Concen: 5.00 ng




RT: 21.75 min Scan# 1375


Delta R.T. 0.00 min

Lab File: BE092295.D

Acq: 29 Aug 2016 18:55

Tgt Ion:	240	Resp:	69105
Ion Ratio	Lower	Upper	
240	100		
120	4.0	1.2	1.8#
236	29.2	17.1	25.7#









#25
Pvrene
Concen: 2.85 ng
RT: 19.97 min Scan# 1287
Delta R.T. -0.02 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Instrument :
BNA_E
ClientSampled :
PB93078BS

Tot Ion	Ratio	Resp Lower	Upper
202	100		
200	5.4	4.2	6.4
203	18.0	14.0	21.0




Manual Integrations
APPROVED


sohil
8/30/2016 4:48:24 PM



#26
Terphenyl-d14
Concen: 56.51 ng
RT: 20.19 min Scan# 1300
Delta R.T. 0.00 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Resp Lower	Upper
244	100		
212	7.4	10.2	15.2#
122	1.2	15.9	23.9#









#27
Benzo(a)anthracene
Concen: 2.71 ng m
RT: 21.72 min Scan# 1374
Delta R.T. 0.00 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Instrument :
BNA_E
ClientSampleId :
PB93078BS

Tot Ion	Ratio	Lower	Upper
228	100		
226	28.6	15.9	23.9#
229	15.7	22.4	33.6#




Manual Integrations
APPROVED


sohil
8/30/2016 4:48:24 PM



#28
Chrysene
Concen: 2.50 ng m
RT: 21.79 min Scan# 1377
Delta R.T. 0.00 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Lower	Upper
228	100		
226	18.8	18.3	27.5
229	28.7	18.6	27.8#









#29
 Bis(2-ethylhexyl)phthalate
 Concen: 2.85 ng
 RT: 21.66 min Scan# 1371
 Delta R.T. 0.00 min
 Lab File: BE092295.D
 Acq: 29 Aug 2016 18:55

Instrument :
 BNA_E
 ClientSampleId :
 PB93078BS

Tot Ion	Ratio	Lower	Upper
149	100		
167	30.4	22.1	33.1
279	2.5	0.0	0.0




Manual Integrations
 APPROVED


sohil
 8/30/2016 4:48:24 PM



#30
 Indeno(1,2,3-cd)pyrene
 Concen: 3.11 ng
 RT: 26.62 min Scan# 1717
 Delta R.T. -0.02 min
 Lab File: BE092295.D
 Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Lower	Upper
276	100		
138	25.6	21.3	31.9
277	25.0	19.7	29.5









#31
Pervlene-d12
Concen: 5.00 ng
RT: 24.12 min Scan# 1547
Delta R.T. 0.00 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Instrument :
BNA_E
ClientSampleId :
PB93078BS

Tot Ion	Ratio	Lower	Upper
264	100		
260	24.7	19.5	29.3
265	22.1	18.2	27.4




Manual Integrations
APPROVED


sohil
8/30/2016 4:48:24 PM



#32
Benzo(b)fluoranthene
Concen: 2.80 ng
RT: 23.38 min Scan# 1483
Delta R.T. -0.01 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.5	19.0	28.4
125	10.1	10.3	15.5#









#33
Benzo(k)fluoranthene
Concen: 2.67 ng
RT: 23.43 min Scan# 1487
Delta R.T. -0.01 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Instrument :
BNA_E
ClientSampleId :
PB93078BS

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.3	19.0	28.6
125	10.3	10.6	15.8




Manual Integrations
APPROVED


sohil
8/30/2016 4:48:24 PM



#34
Benzo(a)pyrene
Concen: 2.79 ng
RT: 24.01 min Scan# 1537
Delta R.T. 0.00 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.0	20.3	30.5
125	11.0	11.8	17.6









#35
Dibenzo(a,h)anthracene
Concen: 3.31 ng
RT: 26.63 min Scan# 1718
Delta R.T. -0.03 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Instrument :
BNA_E
ClientSampleId :
PB93078BS

Tot Ion	Ratio	Lower	Upper
278	100		
139	16.2	16.2	24.2
279	23.4	21.4	32.2

Manual Integrations
APPROVED

sohil
8/30/2016 4:48:24 PM



#36
Benzo(g,h,i)perylene
Concen: 3.15 ng
RT: 27.38 min Scan# 1762
Delta R.T. -0.02 min
Lab File: BE092295.D
Acq: 29 Aug 2016 18:55

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.8	21.4	32.2
138	21.2	19.4	29.2

