

Data Path : Z:\HPCHEM1\BNA_E\Data\BE083116\
 Data File : BE092319.D
 Acq On : 31 Aug 2016 15:43
 Operator : UM/SJ
 Sample : MDL-06-S
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MDL-06-S

Quant Time: Aug 31 16:17:59 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE082516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 26 11:18:03 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	5094	0.40	ng/ul	0.00
2) Naphthalene-d8	10.49	136	27228	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.33	164	16223	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	35454	0.40	ng/ul	0.00
16) Chrysene-d12	21.24	240	37465	0.40	ng/ul	0.00
20) Perylene-d12	23.63	264	36176	0.40	ng/ul	0.00

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.08	152	9804	0.27	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	23227	0.26	ng/ul	0.00

Target Compounds				Qvalue		
3) Naphthalene	10.54	128	3548	0.06	ng/ul	94
5) 2-Methylnaphthalene	12.16	142	2331	0.06	ng/ul	99
7) Acenaphthylene	14.05	152	3932	0.06	ng/ul	96
8) Acenaphthene	14.39	153	2670	0.05	ng/ul	97
9) Fluorene	15.39	166	2933	0.05	ng/ul	98
11) Pentachlorophenol	16.74	266	227	0.05	ng/ul	94
12) Phenanthrene	17.10	178	5030	0.06	ng/ul	96
13) Anthracene	17.20	178	4208	0.05	ng/ul#	92
15) Fluoranthene	19.12	202	5611	0.05	ng/ul	94
17) Pyrene	19.49	202	5798	0.07	ng/ul#	100
18) Benzo(a)anthracene	21.22	228	5204	0.06	ng/ul	95
19) Chrysene	21.27	228	5760	0.06	ng/ul	95
21) Benzo(b)fluoranthene	22.90	252	4967	0.06	ng/ul#	71
22) Benzo(k)fluoranthene	22.95	252	5844	0.06	ng/ul#	66
23) Benzo(a)pyrene	23.52	252	5622	0.06	ng/ul#	61
24) Indeno(1,2,3-cd)pyrene	26.14	276	5920	0.06	ng/ul#	100
25) Dibenzo(a,h)anthracene	26.16	278	4757	0.06	ng/ul#	70
26) Benzo(g,h,i)perylene	26.90	276	5206	0.06	ng/ul#	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\Data\BE083116\
Data File : BE092319.D
Acq On : 31 Aug 2016 15:43
Operator : UM/SJ
Sample : MDL-06-S
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MDL-06-S

Quant Time: Aug 31 16:17:59 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE082516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 26 11:18:03 2016
Response via : Initial Calibration

