

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE083118\
 Data File : BE097407.D
 Acq On : 31 Aug 2018 17:47
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.423

Quant Time: Aug 31 18:25:33 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE082918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 31 18:03:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
3	Naphthalene	0.975	0.983	-0.8	98	0.00
4 SURR	2-Methylnaphthalene-d10	0.622	0.649	-4.3	100	0.00
5	2-Methylnaphthalene	0.616	0.648	-5.2	102	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
7	Acenaphthylene	1.457	1.535	-5.4	111	0.00
8 C	Acenaphthene	1.253	1.268	-1.2	106	0.00
9	Fluorene	1.435	1.490	-3.8	108	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
11	Pentachlorophenol	0.039	0.045	-15.4	142	0.00
12	Phenanthrene	1.030	1.050	-1.9	104	0.00
13	Anthracene	0.866	0.920	-6.2	110	0.00
14 SURR	Fluoranthene-d10	1.319	1.321	-0.2	105	0.00
15 C	Fluoranthene	1.323	1.314	0.7	103	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
17	Pyrene	1.076	1.082	-0.6	101	0.00
18	Benzo(a)anthracene	1.028	1.106	-7.6	110	0.00
19	Chrysene	1.184	1.201	-1.4	102	0.00
20 I	Perylene-d12	1.000	1.000	0.0	102	0.00
21	Benzo(b)fluoranthene	1.071	1.128	-5.3	114	0.00
22	Benzo(k)fluoranthene	1.195	1.150	3.8	98	0.00
23 C	Benzo(a)pyrene	0.991	1.056	-6.6	114	0.00
24	Indeno(1,2,3-cd)pyrene	1.284	1.353	-5.4	108	0.00
25	Dibenzo(a,h)anthracene	1.088	1.135	-4.3	107	0.00
26	Benzo(g,h,i)perylene	1.095	1.146	-4.7	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0