

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090115\
 Data File : BE090702.D
 Acq On : 2 Sep 2015 13:43
 Operator : UM/IZ
 Sample : G3494-02
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 082815-D506-E-U-M

Quant Time: Sep 02 14:43:40 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	39044	5.00	ng	0.00
6) Naphthalene-d8	10.31	136	171193	5.00	ng	0.00
12) Acenaphthene-d10	14.17	164	92822	5.00	ng	0.00
18) Phenanthrene-d10	16.91	188	240241	5.00	ng	0.00
25) Chrysene-d12	21.07	240	299117	5.00	ng	0.00
32) Perylene-d12	23.36	264	275693	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.18	112	32641	4.30	ng	0.00
5) Phenol-d6	6.75	99	24514	2.25	ng	0.00
7) Nitrobenzene-d5	8.71	82	50764	4.49	ng	0.00
13) 2,4,6-Tribromophenol	15.67	330	30263	11.05	ng	0.00
14) 2-Fluorobiphenyl	12.79	172	110204	4.29	ng	0.00
27) Terphenyl-d14	19.53	244	227378	6.87	ng	0.00
Target Compounds						
30) Bis(2-ethylhexyl)phthalate	21.00	149	1855	0.17	ng	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090115\
Data File : BE090702.D
Acq On : 2 Sep 2015 13:43
Operator : UM/IZ
Sample : G3494-02
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
082815-D506-E-U-M

Quant Time: Sep 02 14:43:40 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 02 11:40:50 2015
Response via : Initial Calibration

