

Data Path : Z:\HPCHEM1\BNA_E\Data\BE090115\
 Data File : BE090677.D
 Acq On : 1 Sep 2015 21:45
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE090115

Quant Time: Sep 02 11:46:04 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.814	0.690	15.2	76	0.00
3	n-Nitrosodimethylamine	1.033	0.810	21.6	72	0.00
4 S	2-Fluorophenol	1.039	0.822	20.9	73	0.00
5 S	Phenol-d6	1.392	1.103	20.8	73	0.01
6 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
7 S	Nitrobenzene-d5	0.330	0.269	18.5	71	0.00
8	Nitrobenzene	0.365	0.295	19.2	74	0.00
9	Naphthalene	1.091	0.960	12.0	77	0.00
10	Hexachlorobutadiene	0.171	0.157	8.2	78	0.00
11	2-Methylnaphthalene	0.590	0.501	15.1	77	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
13 S	2,4,6-Tribromophenol	0.143	0.092	35.7#	67	0.02
14 S	2-Fluorobiphenyl	1.385	1.213	12.4	78	0.00
15	Acenaphthylene	7.992	6.742	15.6	77	0.00
16	Acenaphthene	1.335	1.202	10.0	78	0.00
17	Fluorene	1.666	1.464	12.1	78	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
19	4-Bromophenyl-phenylether	0.170	0.148	12.9	81	0.00
20	Hexachlorobenzene	0.928	0.844	9.1	82	0.00
21	Pentachlorophenol	0.054	0.033	38.9#	62	0.02
22	Phenanthrene	1.237	1.088	12.0	79	0.00
23	Anthracene	1.035	0.865	16.4	78	0.00
24	Fluoranthene	1.304	1.036	20.6	76	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
26	Pyrene	1.035	0.793	23.4	75	0.00
27 S	Terphenyl-d14	0.553	0.433	21.7	76	0.00
28	Benzo(a)anthracene	1.156	0.929	19.6	79	0.00
29	Chrysene	1.279	1.119	12.5	83	0.00
30	Bis(2-ethylhexyl)phthalate	0.345	0.195	43.5#	72	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	0.930	21.3	80	0.00
32 I	Perylene-d12	1.000	1.000	0.0	97	0.00
33	Benzo(b)fluoranthene	1.084	0.873	19.5	80	0.00
34	Benzo(k)fluoranthene	1.263	1.057	16.3	82	0.00
35 C	Benzo(a)pyrene	0.970	0.759	21.8#	81	0.00
36	Dibenzo(a,h)anthracene	1.020	0.812	20.4	80	0.00
37	Benzo(g,h,i)perylene	1.116	0.905	18.9	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1