

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090115\
 Data File : BE090688.D
 Acq On : 2 Sep 2015 4:25
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 02 12:25:22 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.814	0.817	-0.4	93	0.00
3	n-Nitrosodimethylamine	1.033	0.942	8.8	87	0.00
4 S	2-Fluorophenol	1.039	0.955	8.1	88	0.00
5 S	Phenol-d6	1.392	1.318	5.3	90	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
7 S	Nitrobenzene-d5	0.330	0.319	3.3	89	0.00
8	Nitrobenzene	0.365	0.339	7.1	89	0.00
9	Naphthalene	1.091	1.092	-0.1	92	0.00
10	Hexachlorobutadiene	0.171	0.174	-1.8	90	0.00
11	2-Methylnaphthalene	0.590	0.580	1.7	93	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
13 S	2,4,6-Tribromophenol	0.143	0.109	23.8	83	0.00
14 S	2-Fluorobiphenyl	1.385	1.366	1.4	92	0.00
15	Acenaphthylene	7.992	7.848	1.8	94	0.00
16	Acenaphthene	1.335	1.362	-2.0	93	0.00
17	Fluorene	1.666	1.627	2.3	91	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
19	4-Bromophenyl-phenylether	0.170	0.168	1.2	91	0.00
20	Hexachlorobenzene	0.928	0.944	-1.7	92	0.00
21	Pentachlorophenol	0.054	0.041	24.1	77	0.00
22	Phenanthrene	1.237	1.253	-1.3	90	0.00
23	Anthracene	1.035	1.010	2.4	90	0.00
24	Fluoranthene	1.304	1.263	3.1	93	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
26	Pyrene	1.035	0.989	4.4	97	0.00
27 S	Terphenyl-d14	0.553	0.520	6.0	94	0.00
28	Benzo(a)anthracene	1.156	1.121	3.0	98	0.00
29	Chrysene	1.279	1.278	0.1	97	0.00
30	Bis(2-ethylhexyl)phthalate	0.345	0.221	35.9#	84	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	1.094	7.4	96	0.00
32 I	Perylene-d12	1.000	1.000	0.0	96	0.00
33	Benzo(b)fluoranthene	1.084	1.054	2.8	96	0.00
34	Benzo(k)fluoranthene	1.263	1.288	-2.0	100	0.00
35 C	Benzo(a)pyrene	0.970	0.934	3.7	99	0.00
36	Dibenzo(a,h)anthracene	1.020	0.976	4.3	96	0.00
37	Benzo(g,h,i)perylene	1.116	1.052	5.7	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0