

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090115\
 Data File : BE090674.D
 Acq On : 1 Sep 2015 19:55
 Operator : UM/IZ
 Sample : SSTDICC002
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Manual Integrations
 APPROVED

MMDadoda
 9/2/2015 4:07:46 PM

Quant Time: Sep 02 01:44:18 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 01:28:15 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	50195	5.00	ng	0.00
6) Naphthalene-d8	10.31	136	215973	5.00	ng	0.00
12) Acenaphthene-d10	14.17	164	109948	5.00	ng	0.00
18) Phenanthrene-d10	16.91	188	265556	5.00	ng	0.00
25) Chrysene-d12	21.07	240	349128	5.00	ng	0.00
32) Perylene-d12	23.36	264	301930	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.19	112	17359	1.25	ng	0.00
5) Phenol-d6	6.75	99	24844	1.22	ng	0.00
7) Nitrobenzene-d5	8.70	82	26350	1.56	ng	0.00
13) 2,4,6-Tribromophenol	15.67	330	4720	1.16	ng	0.00
14) 2-Fluorobiphenyl	12.79	172	54817	1.69	ng	0.00
27) Terphenyl-d14	19.54	244	70179	1.16	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.16	88	13547	2.01	ng	98
3) n-Nitrosodimethylamine	3.45	42	17303	1.96	ng	# 97
8) Nitrobenzene	8.74	77	30386	1.72	ng	99
9) Naphthalene	10.35	128	84418	1.73	ng	99
10) Hexachlorobutadiene	10.65	225	13299	1.83	ng	100
11) 2-Methylnaphthalene	11.98	142	46773	1.50	ng	99
15) Acenaphthylene	13.89	152	315924	1.60	ng	100
16) Acenaphthene	14.23	154	53146	1.76	ng	98
17) Fluorene	15.22	166	66853	1.81	ng	98
19) 4-Bromophenyl-phenylether	16.13	248	16733	1.58	ng	95
20) Hexachlorobenzene	16.23	284	89793	1.86	ng	97
21) Pentachlorophenol	16.60	266	4649m	1.14	ng	
22) Phenanthrene	16.96	178	121546	1.75	ng	99
23) Anthracene	17.05	178	105479	1.68	ng	99
24) Fluoranthene	18.96	202	123393	1.69	ng	99
26) Pyrene	19.32	202	130399	1.35	ng	100
28) Benzo(a)anthracene	21.05	228	144507	1.60	ng	99
29) Chrysene	21.11	228	158508	1.70	ng	99
30) Bis(2-ethylhexyl)phthalate	21.01	149	35429	0.74	ng	100
31) Indeno(1,2,3-cd)pyrene	25.72	276	147718	1.68	ng	99
33) Benzo(b)fluoranthene	22.66	252	124805	1.76	ng	99
34) Benzo(k)fluoranthene	22.71	252	150304	1.87	ng	100
35) Benzo(a)pyrene	23.26	252	108847	1.66	ng	99
36) Dibenzo(a,h)anthracene	25.75	278	118493	1.84	ng	99
37) Benzo(g,h,i)perylene	26.45	276	123099	1.76	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090115\
Data File : BE090674.D
Acq On : 1 Sep 2015 19:55
Operator : UM/IZ
Sample : SSTDICC002
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Manual Integrations
APPROVED

MM Dadoda
9/2/2015 4:07:46 PM

Quant Time: Sep 02 01:44:18 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 02 01:28:15 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.55 min Scan# 838

Delta R.T. -0.00 min

Lab File: BE090674.D


Acq: 1 Sep 2015 19:55

Instrument :

BNA_E

ClientSampleId :

SSTDIC002



Tgt Ion: 152 Resp: 50195

Ion Ratio Lower Upper




152 100

150 157.9 122.2 183.4

115 67.9 51.0 76.4

Manual Integrations
APPROVED

MMDadoda
9/2/2015 4:07:46 PM



#2

1,4-Dioxane

Concen: 2.01 ng

RT: 3.16 min Scan# 83

Delta R.T. -0.00 min

Lab File: BE090674.D

Acq: 1 Sep 2015 19:55




Tgt Ion: 88 Resp: 13547

Ion Ratio Lower Upper

88 100

58 88.4 68.4 102.6

43 33.1 26.9 40.3





#3





n-Nitrosodimethylamine
Concen: 1.96 ng
RT: 3.45 min Scan# 126
Delta R.T. -0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Manual Integrations
APPROVED

MMDadoda
9/2/2015 4:07:46 PM




Tgt Ion	Ratio	Lower	Upper
42	100		
74	106.3	83.7	125.5
44	7.5	10.0	15.0#



#4

2-Fluorophenol
Concen: 1.25 ng
RT: 5.19 min Scan# 382
Delta R.T. -0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55

Tgt Ion	Ratio	Lower	Upper
112	100		
64	71.5	57.3	85.9
63	37.0	29.2	43.8





#5

Phenol-d6

Concen: 1.22 ng

RT: 6.75 min Scan# 661

Delta R.T. -0.00 min

Lab File: BE090674.D

Acq: 1 Sep 2015 19:55

Instrument :

BNA_E


ClientSampleId :

SSTDICC002




Tgt Ion:	99	Resp:	24844
Ion Ratio	Lower	Upper	
99	100		
42	22.1	16.8	25.2
71	34.6	28.3	42.5

Manual Integrations
APPROVED

MMDadoda
9/2/2015 4:07:46 PM



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

Naphthalene-d8

Concen: 5.00 ng

RT: 10.31 min Scan# 1367




Delta R.T. 0.00 min

Lab File: BE090674.D

Acq: 1 Sep 2015 19:55

Tgt Ion:	136	Resp:	215973
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.7	13.1
54	10.6	9.0	13.6
68	7.7	6.3	9.5

Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BE0
Ion 68.00 (67.70 to 68.70): BE0





#7

Nitrobenzene-d5

Concen: 1.56 ng

RT: 8.70 min Scan# 1086

Delta R.T. -0.01 min

Lab File: BE090674.D

Acq: 1 Sep 2015 19:55

Instrument :

BNA_E

ClientSampleId :





SSTDICC002

Manual Integrations
APPROVED

MMDadoda

9/2/2015 4:07:46 PM

Tgt Ion:	82	Resp:	26350
Ion	Ratio	Lower	Upper
82	100		
128	30.4	25.0	37.4
54	60.0	46.3	69.5



#8

Nitrobenzene

Concen: 1.72 ng




RT: 8.74 min Scan# 1095

Delta R.T. -0.01 min

Lab File: BE090674.D

Acq: 1 Sep 2015 19:55

Tgt Ion:	77	Resp:	30386
Ion	Ratio	Lower	Upper
77	100		
123	31.6	25.1	37.7
65	11.9	9.9	14.9





#9

Naphthalene

Concen: 1.73 ng

RT: 10.35 min Scan# 1373

Delta R.T. -0.01 min

Lab File: BE090674.D


Acq: 1 Sep 2015 19:55

Instrument :

BNA_E

ClientSampleId :




SSTDICC002



Tgt Ion:	128	Resp:	84418
Ion Ratio	Lower	Upper	
128	100		
129	10.9	9.0	13.6
127	13.3	10.6	15.8

Manual Integrations
APPROVED

MMDadoda
9/2/2015 4:07:46 PM



#10

Hexachlorobutadiene

Concen: 1.83 ng




RT: 10.65 min Scan# 1412


Delta R.T. 0.00 min

Lab File: BE090674.D

Acq: 1 Sep 2015 19:55

Tgt Ion:	225	Resp:	13299
Ion Ratio	Lower	Upper	
225	100		
223	63.3	50.6	76.0
227	64.3	51.0	76.6









#11
2-Methylnaphthalene
Concen: 1.50 ng
RT: 11.98 min Scan# 1644
Delta R.T. -0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Tgt Ion:142 Resp: 46773
Ion Ratio Lower Upper
142 100
141 90.2 71.4 107.2
115 38.1 31.0 46.4




Manual Integrations
APPROVED


MMDadoda
9/2/2015 4:07:46 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.17 min Scan# 2070
Delta R.T. 0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55

Tgt Ion:164 Resp: 109948
Ion Ratio Lower Upper
164 100
162 107.3 86.8 130.2
160 63.3 53.9 80.9









#13
 2,4,6-Tribromophenol
 Concen: 1.16 ng
 RT: 15.67 min Scan# 2204
 Delta R.T. -0.00 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Tgt Ion:330 Resp: 4720
 Ion Ratio Lower Upper
 330 100
 332 99.9 74.9 112.3
 141 186.3 145.2 217.8




Manual Integrations
 APPROVED

MMDadoda
 9/2/2015 4:07:46 PM



#14
 2-Fluorobiphenyl
 Concen: 1.69 ng
 RT: 12.79 min Scan# 1822
 Delta R.T. -0.00 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55

Tgt Ion:172 Resp: 54817
 Ion Ratio Lower Upper
 172 100
 171 34.5 27.8 41.8
 170 24.5 19.8 29.6





#15

Acenaphthylene

Concen: 1.60 ng

RT: 13.89 min Scan# 2027

Delta R.T. 0.00 min

Lab File: BE090674.D

Acq: 1 Sep 2015 19:55

Instrument :

BNA_E





ClientSampleId :

SSTDICC002

Tgt Ion:	152	Resp:	315924
Ion Ratio	Lower	Upper	
152	100		
151	4.9	3.9	5.9
153	12.9	10.3	15.5

Manual Integrations
APPROVED

MMDadoda
9/2/2015 4:07:46 PM



#16

Acenaphthene

Concen: 1.76 ng




RT: 14.23 min Scan# 2080


Delta R.T. 0.00 min

Lab File: BE090674.D

Acq: 1 Sep 2015 19:55

Tgt Ion:	154	Resp:	53146
Ion Ratio	Lower	Upper	
154	100		
153	443.2	354.5	531.7
152	274.3	227.8	341.6









#17
Fluorene
Concen: 1.81 ng
RT: 15.22 min Scan# 2178
Delta R.T. -0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Tgt Ion	Ratio	Resp	Lower	Upper
166	100	66853		
165	101.4	82.7	124.1	
167	14.1	10.7	16.1	




Manual Integrations
APPROVED



MMDadoda
9/2/2015 4:07:46 PM



#18
Phenanthrene-d10
Concen: 5.00 ng
RT: 16.91 min Scan# 2275
Delta R.T. -0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55

Tgt Ion	Ratio	Resp	Lower	Upper
188	100	265556		
94	14.6	10.3	15.5	
80	15.9	11.0	16.6	









#19
4-Bromophenyl-phenylether
Concen: 1.58 ng
RT: 16.13 min Scan# 2230
Delta R.T. -0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Manual Integrations
APPROVED


MMDadoda
9/2/2015 4:07:46 PM




Tgt Ion	Ratio	Resp	Lower	Upper
248	100	16733		
250	95.2	76.7	115.1	
141	353.1	292.6	439.0	



#20
Hexachlorobenzene
Concen: 1.86 ng
RT: 16.23 min Scan# 2236
Delta R.T. -0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55

Tgt Ion	Ratio	Resp	Lower	Upper
284	100	89793		
142	40.0	35.0	52.4	
249	31.7	25.5	38.3	









#21
 Pentachlorophenol
 Concen: 1.14 ng m
 RT: 16.60 min Scan# 2257
 Delta R.T. -0.00 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55

Tgt Ion	Ratio	Lower	Upper
266	100		
268	68.5	49.6	74.4
264	61.2	53.9	80.9

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002


Manual Integrations
 APPROVED

MMDadoda
 9/2/2015 4:07:46 PM



#22
 Phenanthrene
 Concen: 1.75 ng
 RT: 16.96 min Scan# 2278
 Delta R.T. -0.00 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.6	23.4
179	15.5	12.8	19.2









#23
 Anthracene
 Concen: 1.68 ng
 RT: 17.05 min Scan# 2283
 Delta R.T. -0.00 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Manual Integrations
 APPROVED




MMDadoda
 9/2/2015 4:07:46 PM

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.0	22.6
179	15.1	12.3	18.5



#24
 Fluoranthene
 Concen: 1.69 ng
 RT: 18.96 min Scan# 2501
 Delta R.T. -0.01 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55

Tgt Ion	Ratio	Lower	Upper
202	100		
101	13.8	11.0	16.6
203	17.0	13.9	20.9









#25
Chrysene-d12
Concen: 5.00 ng
RT: 21.07 min Scan# 2861
Delta R.T. -0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Tgt Ion	Ratio	Resp	Lower	Upper
240	100	349128		
120	12.2	10.1	15.1	
236	26.7	21.9	32.9	




Manual Integrations
APPROVED

MMDadoda
9/2/2015 4:07:46 PM



#26
Pyrene
Concen: 1.35 ng
RT: 19.32 min Scan# 2569
Delta R.T. -0.01 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	130399		
200	21.1	16.7	25.1	
203	17.3	13.8	20.6	









#27
 Terphenyl-d14
 Concen: 1.16 ng
 RT: 19.54 min Scan# 2610
 Delta R.T. -0.00 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Tgt Ion	Ratio	Lower	Upper
244	100		
212	8.2	6.9	10.3
122	14.1	12.9	19.3




Manual Integrations
 APPROVED


MMDadoda
 9/2/2015 4:07:46 PM



#28
 Benzo(a)anthracene
 Concen: 1.60 ng
 RT: 21.05 min Scan# 2858
 Delta R.T. -0.01 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.7	21.6	32.4
229	19.2	15.9	23.9





#29

Chrysene

Concen: 1.70 ng

RT: 21.11 min Scan# 2866

Delta R.T. -0.00 min

Lab File: BE090674.D


Acq: 1 Sep 2015 19:55

Instrument :

BNA_E

ClientSampleId :




SSTDICC002



Tgt Ion:	228	Resp:	158508
Ion Ratio	Lower	Upper	
228	100		
226	30.6	24.2	36.4
229	19.6	15.3	22.9

Manual Integrations
APPROVED

MMDadoda
9/2/2015 4:07:46 PM



#30

Bis(2-ethylhexyl)phthalate

Concen: 0.74 ng




RT: 21.01 min Scan# 2851


Delta R.T. -0.00 min

Lab File: BE090674.D

Acq: 1 Sep 2015 19:55


Tgt Ion:	149	Resp:	35429
Ion Ratio	Lower	Upper	
149	100		
167	25.0	20.1	30.1
279	2.9	2.3	3.5





#31
 Indeno(1,2,3-cd)pyrene
 Concen: 1.68 ng
 RT: 25.72 min Scan# 3814
 Delta R.T. -0.00 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002




Tgt Ion	Ratio	Resp	Lower	Upper
276	100	147718		
138	29.6	23.3	34.9	
277	24.5	19.8	29.6	

Manual Integrations
 APPROVED



MMDadoda
 9/2/2015 4:07:46 PM




#32
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.36 min Scan# 3299
 Delta R.T. -0.00 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55




Tgt Ion	Ratio	Resp	Lower	Upper
264	100	301930		
260	24.8	19.7	29.5	
265	21.6	17.5	26.3	





#33
Benzo(b)fluoranthene
Concen: 1.76 ng
RT: 22.66 min Scan# 3146
Delta R.T. -0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55




Instrument :
BNA_E
ClientSampleId :
SSTDICC002




Tgt Ion	Ratio	Resp Lower	Resp Upper
252	100		
253	21.8	17.9	26.9
125	12.9	10.6	16.0

Manual Integrations
APPROVED



MMDadoda
9/2/2015 4:07:46 PM




#34
Benzo(k)fluoranthene
Concen: 1.87 ng
RT: 22.71 min Scan# 3157
Delta R.T. 0.00 min
Lab File: BE090674.D
Acq: 1 Sep 2015 19:55



Tgt Ion	Ratio	Resp Lower	Resp Upper
252	100		
253	22.2	17.9	26.9
125	12.5	10.1	15.1





#35

Benzo(a)pyrene

Concen: 1.66 ng

RT: 23.26 min Scan# 3276

Delta R.T. -0.00 min

Lab File: BE090674.D


Acq: 1 Sep 2015 19:55

Instrument :

BNA_E

ClientSampleId :




SSTDICC002



Tgt Ion:	252	Resp:	108847
Ion Ratio	Lower	Upper	
252	100		
253	21.6	17.5	26.3
125	14.2	11.6	17.4

Manual Integrations
APPROVED

MMDadoda
9/2/2015 4:07:46 PM



#36

Dibenzo(a,h)anthracene

Concen: 1.84 ng




RT: 25.75 min Scan# 3819


Delta R.T. -0.00 min

Lab File: BE090674.D

Acq: 1 Sep 2015 19:55

Tgt Ion:	278	Resp:	118493
Ion Ratio	Lower	Upper	
278	100		
139	18.5	15.4	23.0
279	24.2	18.8	28.2






#37
 Benzo(g,h,i)perylene
 Concen: 1.76 ng
 RT: 26.45 min Scan# 3973
 Delta R.T. -0.01 min
 Lab File: BE090674.D
 Acq: 1 Sep 2015 19:55

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Tgt	Ion	276	Resp:	123099
Ion	Ratio	Lower	Upper	
276	100			
277	23.4	18.5	27.7	
138	25.5	20.1	30.1	

Manual Integrations
 APPROVED

MMDadoda
 9/2/2015 4:07:46 PM



Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 277.00 (276.70 to 277.70): E
 Ion 138.00 (137.70 to 138.70): E

