


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090115\
Data File : BE090687.D
Acq On : 2 Sep 2015 3:50
Operator : UM/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Sep 02 11:40:50 2015



AutoFind: Scans 2513, 2514, 2515; Background Corrected with Scan 2505

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	56.0	19003	PASS
68	69	0.00	2	0.3	57	PASS
69	198	0.00	100	53.8	18274	PASS
70	69	0.00	2	0.5	100	PASS
127	198	10	80	56.2	19071	PASS
197	198	0.00	2	0.2	80	PASS
198	198	100	100	100.0	33948	PASS
199	198	5	9	6.3	2133	PASS
275	198	10	60	24.8	8410	PASS
365	198	1	100	3.6	1206	PASS
441	442	0.01	24	18.0	5524	PASS
442	198	50	100	90.4	30680	PASS
443	442	15	24	19.6	6014	PASS