

Data Path : Z:\HPCHEM1\BNA_E\Data\BE090117\
 Data File : BE093921.D
 Acq On : 2 Sep 2017 5:17
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 04 01:28:38 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 01 16:31:36 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	133	0.00
2	1,4-Dioxane	0.550	0.544	1.1	140	0.01
3	n-Nitrosodimethylamine	0.612	0.599	2.1	137	0.00
4 S	2-Fluorophenol	1.231	1.176	4.5	129	0.00
5 S	Phenol-d6	1.882	1.759	6.5	131	0.01
6	bis(2-Chloroethyl)ether	1.450	1.409	2.8	133	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	136	0.00
8 S	Nitrobenzene-d5	0.333	0.301	9.6	128	0.00
9	Naphthalene	4.555	3.977	12.7	120	0.00
10	Hexachlorobutadiene	0.161	0.147	8.7	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.550	13.1	120	0.00
12	2-Methylnaphthalene	0.761	0.660	13.3	122	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.107	0.095	11.2	141	0.00
15 S	2-Fluorobiphenyl	1.592	1.379	13.4	122	-0.02
16	Acenaphthylene	1.979	1.745	11.8	129	0.00
17	Acenaphthene	1.335	1.168	12.5	126	0.00
18	Fluorene	1.727	1.483	14.1	124	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	134	0.00
20	4-Bromophenyl-phenylether	0.166	0.137	17.5	110	0.00
21	Hexachlorobenzene	0.144	0.135	6.2	120	0.00
22	Pentachlorophenol	0.063	0.085	-34.9#	224#	-0.03
23	Phenanthrene	1.017	0.824	19.0	111	-0.03
24	Anthracene	1.212	1.038	14.4	124	0.00
25 SURR	Fluoranthene-d10	4.612	4.650	-0.8	129	0.00
26	Fluoranthene	1.393	1.362	2.2	126	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	139	0.00
28	Pyrene	1.346	1.184	12.0	129	0.00
29 S	Terphenyl-d14	0.724	0.639	11.7	129	0.00
30	Benzo(a)anthracene	1.188	1.085	8.7	137	-0.01
31	Chrysene	1.320	1.114	15.6	123	0.00
32	Bis(2-ethylhexyl)phthalate	2.504	2.754	-10.0	171#	0.00
33	Indeno(1,2,3-cd)pyrene	0.901	0.959	-6.4	159#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	147	0.00
35	Benzo(b)fluoranthene	1.277	1.149	10.0	146	0.00
36	Benzo(k)fluoranthene	1.460	1.237	15.3	128	0.00
37 C	Benzo(a)pyrene	1.268	1.140	10.1	142	0.00
38	Dibenzo(a,h)anthracene	0.895	0.934	-4.4	162#	0.00
39	Benzo(g,h,i)perylene	0.975	0.954	2.2	155#	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE090117\
Data File : BE093921.D
Acq On : 2 Sep 2017 5:17
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 04 01:28:38 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 01 16:31:36 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		