


Data Path : Z:\HPCHEM1\BNA_E\Data\BE090117\
 Data File : BE093920.D
 Acq On : 2 Sep 2017 4:41
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE082617.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Sep 01 16:31:36 2017



AutoFind: Scans 2569, 2570, 2571; Background Corrected with Scan 2561

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.7	12514	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	42.7	12512	PASS
70	69	0.00	2	0.5	57	PASS
127	198	10	80	49.4	14475	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	29293	PASS
199	198	5	9	6.9	2011	PASS
275	198	10	60	24.7	7246	PASS
365	198	1	100	2.8	827	PASS
441	442	0.01	24	16.2	3278	PASS
442	198	50	100	69.0	20217	PASS
443	442	15	24	19.2	3889	PASS