

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090215\
 Data File : BE090704.D
 Acq On : 2 Sep 2015 14:55
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 02 23:03:36 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	80	0.00
2	1,4-Dioxane	1.000	1.002	-0.2	75	0.00
3	n-Nitrosodimethylamine	1.000	0.990	1.0	76	0.00
4 S	2-Fluorophenol	1.000	0.995	0.5	78	0.00
5 S	Phenol-d6	1.000	0.945	5.5	79	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	82	0.00
7 S	Nitrobenzene-d5	1.000	0.962	3.8	79	0.00
8	Nitrobenzene	1.000	0.900	10.0	77	0.00
9	Naphthalene	1.000	1.006	-0.6	83	0.00
10	Hexachlorobutadiene	1.000	1.026	-2.6	82	0.00
11	2-Methylnaphthalene	1.000	1.010	-1.0	86	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	83	0.00
13 S	2,4,6-Tribromophenol	1.000	0.975	2.5	84	0.00
14 S	2-Fluorobiphenyl	1.000	1.004	-0.4	85	0.00
15	Acenaphthylene	1.000	1.001	-0.1	87	0.00
16	Acenaphthene	1.000	1.038	-3.8	85	0.00
17	Fluorene	1.000	1.025	-2.5	86	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	85	0.00
19	4-Bromophenyl-phenylether	1.000	0.970	3.0	86	0.00
20	Hexachlorobenzene	1.000	1.044	-4.4	87	0.00
21	Pentachlorophenol	1.000	0.979	2.1	69	0.00
22	Phenanthrene	1.000	1.003	-0.3	86	-0.02
23	Anthracene	1.000	0.981	1.9	87	0.00
24	Fluoranthene	1.000	0.967	3.3	89	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	89	0.00
26	Pyrene	1.000	0.986	1.4	92	0.00
27 S	Terphenyl-d14	1.000	0.949	5.1	88	0.00
28	Benzo(a)anthracene	1.000	1.013	-1.3	91	0.00
29	Chrysene	1.000	1.055	-5.5	90	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.812	18.8	79	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.034	-3.4	99	0.00
32 I	Perylene-d12	5.000	5.000	0.0	91	0.00
33	Benzo(b)fluoranthene	1.000	1.016	-1.6	95	0.00
34	Benzo(k)fluoranthene	1.000	1.010	-1.0	93	0.00
35 C	Benzo(a)pyrene	1.000	0.998	0.2	97	0.00
36	Dibenzo(a,h)anthracene	1.000	0.998	0.2	97	0.00
37	Benzo(g,h,i)perylene	1.000	0.973	2.7	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0