

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090216\
 Data File : BE092350.D
 Acq On : 2 Sep 2016 19:14
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE090216

Quant Time: Sep 05 09:08:43 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	133	0.00
2	1,4-Dioxane	0.400	0.395	1.3	135	0.00
3	n-Nitrosodimethylamine	0.400	0.427	-6.7	170	-0.01
4 S	2-Fluorophenol	0.400	0.388	3.0	152	0.00
5 S	Phenol-d6	0.400	0.376	6.0	162	0.00
6	bis(2-Chloroethyl)ether	0.400	0.419	-4.7	175	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	137	0.00
8 S	Nitrobenzene-d5	0.400	0.375	6.3	163	0.00
9	Naphthalene	0.400	0.417	-4.2	150	0.00
10	Hexachlorobutadiene	0.400	0.422	-5.5	147	0.00
11	2-Methylnaphthalene	0.400	0.410	-2.5	151	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	131	0.00
13 S	2,4,6-Tribromophenol	0.400	0.397	0.8	162	0.00
14 S	2-Fluorobiphenyl	0.400	0.415	-3.7	151	0.00
15	Acenaphthylene	0.400	0.414	-3.5	151	0.00
16	Acenaphthene	0.400	0.416	-4.0	148	0.00
17	Fluorene	0.400	0.415	-3.7	151	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	132	0.02
19	Hexachlorobenzene	0.400	0.407	-1.7	147	0.00
20	Pentachlorophenol	0.400	0.422	-5.5	174	0.00
21	Phenanthrene	0.400	0.393	1.8	142	0.00
22	Anthracene	0.400	0.390	2.5	155	0.02
23	Fluoranthene	0.400	0.416	-4.0	155	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	157	0.00
25	Pyrene	0.400	0.387	3.3	159	0.00
26 S	Terphenyl-d14	0.400	0.373	6.8	155	0.00
27	Benzo(a)anthracene	0.400	0.402	-0.5	161	0.00
28	Chrysene	0.400	0.356	11.0	180	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.320	20.0	148	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.384	4.0	154	0.02
31 I	Perylene-d12	5.000	5.000	0.0	134	0.02
32	Benzo(b)fluoranthene	0.400	0.451	-12.7	171	0.02
33	Benzo(k)fluoranthene	0.400	0.384	4.0	134	0.02
34 C	Benzo(a)pyrene	0.400	0.416	-4.0	156	0.01
35	Dibenzo(a,h)anthracene	0.400	0.410	-2.5	155	0.04
36	Benzo(g,h,i)perylene	0.400	0.406	-1.5	146	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0