

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090216\
 Data File : BE092350.D
 Acq On : 2 Sep 2016 19:14
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE090216

Quant Time: Sep 05 09:08:43 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	133	0.00
2	1,4-Dioxane	0.707	0.738	-4.4	135	0.00
3	n-Nitrosodimethylamine	0.661	0.708	-7.1	170#	-0.01
4 S	2-Fluorophenol	1.189	1.153	3.0	152#	0.00
5 S	Phenol-d6	1.586	1.489	6.1	162#	0.00
6	bis(2-Chloroethyl)ether	1.288	1.340	-4.0	175#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	137	0.00
8 S	Nitrobenzene-d5	0.359	0.336	6.4	163#	0.00
9	Naphthalene	1.120	1.168	-4.3	150	0.00
10	Hexachlorobutadiene	0.175	0.184	-5.1	147	0.00
11	2-Methylnaphthalene	0.733	0.751	-2.5	151#	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.00
13 S	2,4,6-Tribromophenol	0.155	0.154	0.6	162#	0.00
14 S	2-Fluorobiphenyl	1.534	1.591	-3.7	151#	0.00
15	Acenaphthylene	8.606	8.907	-3.5	151#	0.00
16	Acenaphthene	1.393	1.448	-3.9	148	0.00
17	Fluorene	1.749	1.814	-3.7	151#	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	132	0.02
19	Hexachlorobenzene	0.239	0.244	-2.1	147	0.00
20	Pentachlorophenol	0.053	0.047	11.3	174#	0.00
21	Phenanthrene	1.405	1.379	1.9	142	0.00
22	Anthracene	1.250	1.219	2.5	155#	0.02
23	Fluoranthene	6.135	6.380	-4.0	155#	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	157#	0.00
25	Pyrene	4.864	4.709	3.2	159#	0.00
26 S	Terphenyl-d14	0.740	0.690	6.8	155#	0.00
27	Benzo(a)anthracene	5.396	5.427	-0.6	161#	0.00
28	Chrysene	5.570	5.754	-3.3	180#	0.00
29	Bis(2-ethylhexyl)phthalate	0.749	0.585	21.9	148	0.00
30	Indeno(1,2,3-cd)pyrene	5.684	5.461	3.9	154#	0.02
31 I	Perylene-d12	1.000	1.000	0.0	134	0.02
32	Benzo(b)fluoranthene	1.284	1.447	-12.7	171#	0.02
33	Benzo(k)fluoranthene	1.357	1.303	4.0	134	0.02
34 C	Benzo(a)pyrene	1.265	1.314	-3.9	156#	0.01
35	Dibenzo(a,h)anthracene	1.170	1.199	-2.5	155#	0.04
36	Benzo(g,h,i)perylene	4.933	5.008	-1.5	146	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0