

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090216\
 Data File : BE092354.D
 Acq On : 2 Sep 2016 21:44
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 02 23:56:55 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	99	0.00
2	1,4-Dioxane	0.400	0.370	7.5	95	0.00
3	n-Nitrosodimethylamine	0.400	0.369	7.8	106	0.00
4 S	2-Fluorophenol	0.400	0.340	15.0	99	0.00
5 S	Phenol-d6	0.400	0.312	22.0	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.352	12.0	105	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.400	0.315	21.3	102	0.00
9	Naphthalene	0.400	0.378	5.5	101	0.00
10	Hexachlorobutadiene	0.400	0.392	2.0	102	0.00
11	2-Methylnaphthalene	0.400	0.364	9.0	100	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	99	0.00
13 S	2,4,6-Tribromophenol	0.400	0.332	17.0	102	0.00
14 S	2-Fluorobiphenyl	0.400	0.371	7.3	101	0.01
15	Acenaphthylene	0.400	0.369	7.8	101	0.02
16	Acenaphthene	0.400	0.379	5.3	101	0.00
17	Fluorene	0.400	0.368	8.0	100	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	97	0.02
19	Hexachlorobenzene	0.400	0.366	8.5	98	0.00
20	Pentachlorophenol	0.400	0.362	9.5	106	0.00
21	Phenanthrene	0.400	0.361	9.8	97	0.00
22	Anthracene	0.400	0.344	14.0	101	0.02
23	Fluoranthene	0.400	0.367	8.3	101	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	118	0.00
25	Pyrene	0.400	0.338	15.5	104	0.00
26 S	Terphenyl-d14	0.400	0.319	20.3	100	0.02
27	Benzo(a)anthracene	0.400	0.358	10.5	108	0.00
28	Chrysene	0.400	0.263	34.3#	108	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.232	42.0#	69	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.331	17.3	99	0.03
31 I	Perylene-d12	5.000	5.000	0.0	98	0.02
32	Benzo(b)fluoranthene	0.400	0.395	1.3	109	0.02
33	Benzo(k)fluoranthene	0.400	0.368	8.0	93	0.02
34 C	Benzo(a)pyrene	0.400	0.384	4.0	105	0.01
35	Dibenzo(a,h)anthracene	0.400	0.362	9.5	100	0.03
36	Benzo(g,h,i)perylene	0.400	0.366	8.5	96	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0