

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090216\
 Data File : BE092354.D
 Acq On : 2 Sep 2016 21:44
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 02 23:56:55 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.707	0.700	1.0	95	0.00
3	n-Nitrosodimethylamine	0.661	0.590	10.7	106	0.00
4 S	2-Fluorophenol	1.189	1.010	15.1	99	0.00
5 S	Phenol-d6	1.586	1.237	22.0	100	0.00
6	bis(2-Chloroethyl)ether	1.288	1.080	16.1	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.359	0.283	21.2	102	0.00
9	Naphthalene	1.120	1.057	5.6	101	0.00
10	Hexachlorobutadiene	0.175	0.171	2.3	102	0.00
11	2-Methylnaphthalene	0.733	0.667	9.0	100	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
13 S	2,4,6-Tribromophenol	0.155	0.129	16.8	102	0.00
14 S	2-Fluorobiphenyl	1.534	1.423	7.2	101	0.01
15	Acenaphthylene	8.606	7.934	7.8	101	0.02
16	Acenaphthene	1.393	1.319	5.3	101	0.00
17	Fluorene	1.749	1.607	8.1	100	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.02
19	Hexachlorobenzene	0.239	0.219	8.4	98	0.00
20	Pentachlorophenol	0.053	0.039	26.4#	106	0.00
21	Phenanthrene	1.405	1.269	9.7	97	0.00
22	Anthracene	1.250	1.076	13.9	101	0.02
23	Fluoranthene	6.135	5.627	8.3	101	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
25	Pyrene	4.864	4.105	15.6	104	0.00
26 S	Terphenyl-d14	0.740	0.589	20.4	100	0.02
27	Benzo(a)anthracene	5.396	4.829	10.5	108	0.00
28	Chrysene	5.570	4.631	16.9	108	0.00
29	Bis(2-ethylhexyl)phthalate	0.749	0.364	51.4#	69	0.00
30	Indeno(1,2,3-cd)pyrene	5.684	4.708	17.2	99	0.03
31 I	Perylene-d12	1.000	1.000	0.0	98	0.02
32	Benzo(b)fluoranthene	1.284	1.268	1.2	109	0.02
33	Benzo(k)fluoranthene	1.357	1.249	8.0	93	0.02
34 C	Benzo(a)pyrene	1.265	1.215	4.0	105	0.01
35	Dibenzo(a,h)anthracene	1.170	1.058	9.6	100	0.03
36	Benzo(g,h,i)perylene	4.933	4.519	8.4	96	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0