

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090216\
 Data File : BE092343.D
 Acq On : 2 Sep 2016 14:58
 Operator : UM/SJ
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2

Manual Integrations
 APPROVED

sohil
 9/6/2016 1:15:07 PM

Quant Time: Sep 02 16:27:43 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 16:14:41 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.19	152	8915	5.00	ng	0.00
7) Naphthalene-d8	10.95	136	42301	5.00	ng	0.00
12) Acenaphthene-d10	14.81	164	26215	5.00	ng	0.00
18) Phenanthrene-d10	17.56	188	66484	5.00	ng	0.00
24) Chrysene-d12	21.72	240	78452	5.00	ng	-0.02
31) Perylene-d12	24.09	264	89469	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.69	112	381	0.16	ng	0.00
5) Phenol-d6	7.36	99	539	0.16	ng	0.00
8) Nitrobenzene-d5	9.36	82	604	0.20	ng	0.00
13) 2,4,6-Tribromophenol	16.32	330	144	0.14	ng	0.00
14) 2-Fluorobiphenyl	13.46	172	1619	0.20	ng	0.01
26) Terphenyl-d14	20.17	244	2524	0.23	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.42	88	275	0.22	ng	98
3) n-Nitrosodimethylamine	3.80	42	190	0.14	ng	# 2
6) bis(2-Chloroethyl)ether	7.61	93	408	0.16	ng	# 5
9) Naphthalene	11.01	128	1949	0.21	ng	94
10) Hexachlorobutadiene	11.30	225	315	0.21	ng	100
11) 2-Methylnaphthalene	12.66	142	1193	0.19	ng	97
15) Acenaphthylene	14.54	152	8769	0.19	ng	100
16) Acenaphthene	14.88	154	1551	0.21	ng	97
17) Fluorene	15.87	166	1826	0.20	ng	99
19) Hexachlorobenzene	16.89	284	592	0.20	ng	# 64
20) Pentachlorophenol	17.26	266	90	0.10	ng	88
21) Phenanthrene	17.60	178	3642	0.23	ng	# 99
22) Anthracene	17.69	178	2838	0.18	ng	100
23) Fluoranthene	19.61	202	15214	0.19	ng	99
25) Pyrene	19.97	202	16120	0.23	ng	99
27) Benzo(a)anthracene	21.70	228	20304	0.23	ng	# 82
28) Chrysene	21.77	228	15862m	0.22	ng	
29) Bis(2-ethylhexyl)phthalate	21.63	149	2013	0.24	ng	# 69
30) Indeno(1,2,3-cd)pyrene	26.58	276	19540	0.23	ng	99
32) Benzo(b)fluoranthene	23.36	252	4403	0.18	ng	96
33) Benzo(k)fluoranthene	23.41	252	4961	0.21	ng	97
34) Benzo(a)pyrene	23.98	252	4487	0.20	ng	96
35) Dibenzo(a,h)anthracene	26.60	278	3997	0.18	ng	# 94
36) Benzo(g,h,i)perylene	27.35	276	17474	0.20	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090216\
Data File : BE092343.D
Acq On : 2 Sep 2016 14:58
Operator : UM/SJ
Sample : SSTDIC0.2
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleID :
SSTDIC0.2

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:07 PM

Quant Time: Sep 02 16:27:43 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 16:14:41 2016
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 8.19 min Scan# 577
Delta R.T. -0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2

Tgt Ion	Ratio	Lower	Upper
152	100		
150	132.3	106.0	159.0
115	39.2	31.2	46.8

Manual Integrations
APPROVED




sohil
9/6/2016 1:15:07 PM




#2

1,4-Dioxane
Concen: 0.22 ng
RT: 3.42 min Scan# 72
Delta R.T. 0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Tgt Ion	Ratio	Lower	Upper
88	100		
58	77.5	63.6	95.4
43	34.9	28.0	42.0





#3

n-Nitrosodimethylamine

Concen: 0.14 ng

RT: 3.80 min Scan# 105

Delta R.T. 0.00 min

Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Instrument :

BNA_E





ClientSampleId :

SSTDICC0.2

Tgt Ion:	42	Resp:	190
Ion	Ratio	Lower	Upper
42	100		
74	0.0	86.0	129.0#
44	4.2	5.1	7.7#

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:07 PM



#4

2-Fluorophenol

Concen: 0.16 ng




RT: 5.69 min Scan# 366


Delta R.T. -0.00 min

Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Tgt Ion:	112	Resp:	381
Ion	Ratio	Lower	Upper
112	100		
64	66.7	53.5	80.3
63	32.3	27.9	41.9





#5

Phenol-d6

Concen: 0.16 ng

RT: 7.36 min Scan# 541

Delta R.T. -0.00 min

Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Instrument :

BNA_E

ClientSampleId :





SSTDICC0.2

Tgt Ion: 99 Resp: 539

Ion	Ratio	Lower	Upper
99	100		
42	23.0	0.0	0.0#
71	35.6	30.6	45.8

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:07 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.16 ng

RT: 7.61 min Scan# 552




Delta R.T. -0.00 min


Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Tgt Ion: 93 Resp: 408

Ion	Ratio	Lower	Upper
93	100		
63	0.0	71.6	107.4#
95	34.3	0.0	0.0#





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.95 min Scan# 708

Delta R.T. 0.00 min

Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Instrument :

BNA_E


ClientSampleId :

SSTDIC0.2

Tgt Ion	Ratio	Resp	Lower	Upper
136	100	42301		
137	10.6	8.2	12.4	
54	9.6	9.6	14.4	
68	7.0	6.7	10.1	

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:07 PM




Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E


Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0

Time-->



Time-->



#8

Nitrobenzene-d5

Concen: 0.20 ng


RT: 9.36 min Scan# 628

Delta R.T. -0.00 min

Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	604		
128	40.8	39.0	58.4	
54	72.5	44.8	67.2	




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0

Time-->



Time-->

Time-->



#9

Naphthalene

Concen: 0.21 ng

RT: 11.01 min Scan# 711

Delta R.T. 0.00 min

Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.2

Tgt Ion:	128	Resp:	1949
Ion Ratio	Lower	Upper	
128	100		
129	16.7	11.2	16.8
127	16.8	11.6	17.4

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:07 PM



#10

Hexachlorobutadiene

Concen: 0.21 ng




RT: 11.30 min Scan# 726

Delta R.T. 0.00 min

Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Tgt Ion:	225	Resp:	315
Ion Ratio	Lower	Upper	
225	100		
223	63.2	50.6	76.0
227	64.1	51.4	77.0









#11
2-Methylnaphthalene
Concen: 0.19 ng
RT: 12.66 min Scan# 817
Delta R.T. 0.01 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.2

Tgt Ion:142 Resp: 1193
Ion Ratio Lower Upper
142 100
141 88.0 73.7 110.5
115 41.8 34.0 51.0




Manual Integrations
APPROVED


sohil
9/6/2016 1:15:07 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.81 min Scan# 979
Delta R.T. -0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Tgt Ion:164 Resp: 26215
Ion Ratio Lower Upper
164 100
162 111.2 71.4 107.0#
160 53.7 27.4 41.2#






#13
 2,4,6-Tribromophenol
 Concen: 0.14 ng
 RT: 16.32 min Scan# 1088
 Delta R.T. -0.00 min
 Lab File: BE092343.D
 Acq: 2 Sep 2016 14:58

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2


Tgt Ion: 330 Resp: 144
 Ion Ratio Lower Upper
 330 100
 332 95.8 75.4 113.0
 141 45.1 31.0 46.6

Manual Integrations
 APPROVED



sohil
 9/6/2016 1:15:07 PM



Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E




Time-->




#14
 2-Fluorobiphenyl
 Concen: 0.20 ng
 RT: 13.46 min Scan# 886
 Delta R.T. 0.01 min
 Lab File: BE092343.D
 Acq: 2 Sep 2016 14:58


Tgt Ion: 172 Resp: 1619
 Ion Ratio Lower Upper
 172 100
 171 37.3 30.1 45.1
 170 25.9 21.8 32.6




Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E



Time-->









#15
Acenaphthylene
Concen: 0.19 ng
RT: 14.54 min Scan# 963
Delta R.T. 0.02 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2

Tgt Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.9	5.9
153	13.0	10.4	15.6




Manual Integrations
APPROVED


sohil
9/6/2016 1:15:07 PM



#16
Acenaphthene
Concen: 0.21 ng
RT: 14.88 min Scan# 983
Delta R.T. -0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Tgt Ion	Ratio	Lower	Upper
154	100		
153	428.4	350.8	526.2
152	213.6	171.1	256.7









#17
Fluorene
 Concen: 0.20 ng
 RT: 15.87 min Scan# 1049
 Delta R.T. -0.00 min
 Lab File: BE092343.D
 Acq: 2 Sep 2016 14:58

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2

Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.1	78.2	117.4
167	13.2	11.0	16.4




Manual Integrations
 APPROVED

sohil
 9/6/2016 1:15:07 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.56 min Scan# 1152
 Delta R.T. -0.00 min
 Lab File: BE092343.D
 Acq: 2 Sep 2016 14:58

Tgt Ion	Ratio	Lower	Upper
188	100		
94	11.9	3.2	4.8#
80	11.9	2.6	3.8#






#19
Hexachlorobenzene
Concen: 0.20 ng
RT: 16.89 min Scan# 1123
Delta R.T. -0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion	Ratio	Lower	Upper
284	100		
142	0.0	29.1	43.7#
249	29.4	27.9	41.9

Manual Integrations
APPROVED



sohil
9/6/2016 1:15:07 PM



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E




Time-->




#20
Pentachlorophenol
Concen: 0.10 ng
RT: 17.26 min Scan# 1139
Delta R.T. -0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58


Tgt Ion	Ratio	Lower	Upper
266	100		
268	86.1	62.5	93.7
264	83.3	57.2	85.8




Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E



Time-->









#21
Phenanthrene
Concen: 0.23 ng
RT: 17.60 min Scan# 1154
Delta R.T. -0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.8	23.8
179	16.3	0.0	0.0#




Manual Integrations
APPROVED


sohil
9/6/2016 1:15:07 PM



#22
Anthracene
Concen: 0.18 ng
RT: 17.69 min Scan# 1158
Delta R.T. -0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.0	22.4
179	15.2	12.2	18.2









#23
 Fluoranthene
 Concen: 0.19 ng
 RT: 19.61 min Scan# 1265
 Delta R.T. 0.00 min
 Lab File: BE092343.D
 Acq: 2 Sep 2016 14:58

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2

Tgt Ion	Ratio	Lower	Upper
202	100		
101	2.8	2.2	3.4
203	17.6	13.7	20.5




Manual Integrations
 APPROVED

sohil
 9/6/2016 1:15:07 PM



#24
 Chrysene-d12
 Concen: 5.00 ng
 RT: 21.72 min Scan# 1373
 Delta R.T. -0.02 min
 Lab File: BE092343.D
 Acq: 2 Sep 2016 14:58

Tgt Ion	Ratio	Lower	Upper
240	100		
120	14.1	2.6	4.0#
236	34.5	24.6	37.0









#25
 Pyrene
 Concen: 0.23 ng
 RT: 19.97 min Scan# 1286
 Delta R.T. 0.00 min
 Lab File: BE092343.D
 Acq: 2 Sep 2016 14:58

Instrument :
 BNA_E
 ClientSampled :
 SSTDICC0.2

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	16120		
200	5.3	4.2	6.4	
203	17.9	13.9	20.9	




Manual Integrations
 APPROVED

sohil
 9/6/2016 1:15:07 PM



#26
 Terphenyl-d14
 Concen: 0.23 ng
 RT: 20.17 min Scan# 1298
 Delta R.T. 0.00 min
 Lab File: BE092343.D
 Acq: 2 Sep 2016 14:58

Tgt Ion	Ratio	Resp	Lower	Upper
244	100	2524		
212	11.0	6.7	10.1#	
122	11.3	3.6	5.4#	









#27
Benzo(a)anthracene
Concen: 0.23 ng
RT: 21.70 min Scan# 1372
Delta R.T. -0.02 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.2

Tgt Ion	Ratio	Lower	Upper
228	100		
226	42.2	23.6	35.4#
229	13.5	13.3	19.9




Manual Integrations
APPROVED


sohil
9/6/2016 1:15:07 PM



#28
Chrysene
Concen: 0.22 ng m
RT: 21.77 min Scan# 1375
Delta R.T. -0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Tgt Ion	Ratio	Lower	Upper
228	100		
226	23.0	36.3	54.5#
229	24.8	10.6	16.0#





#29

Bis(2-ethylhexyl)phthalate

Concen: 0.24 ng

RT: 21.63 min Scan# 1369

Delta R.T. -0.00 min

Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Instrument :

BNA_E


ClientSampleId :

SSTDIC0.2




Tgt Ion:	149	Resp:	2013
Ion Ratio	Lower	Upper	
149	100		
167	28.2	16.0	24.0#
279	0.0	16.0	24.0#

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:07 PM



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E



#30

Indeno(1,2,3-cd)pyrene

Concen: 0.23 ng

RT: 26.58 min Scan# 1714




Delta R.T. 0.00 min


Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Tgt Ion:	276	Resp:	19540
Ion Ratio	Lower	Upper	
276	100		
138	26.7	20.3	30.5
277	24.6	19.7	29.5

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#31
Perylene-d12
 Concen: 5.00 ng
 RT: 24.09 min Scan# 1543
 Delta R.T. 0.00 min
 Lab File: BE092343.D
 Acq: 2 Sep 2016 14:58

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.4	20.1	30.1
265	22.0	17.9	26.9




Manual Integrations
 APPROVED

sohil
 9/6/2016 1:15:07 PM



#32
Benzo(b)fluoranthene
 Concen: 0.18 ng
 RT: 23.36 min Scan# 1480
 Delta R.T. 0.00 min
 Lab File: BE092343.D
 Acq: 2 Sep 2016 14:58

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.7	18.7	28.1
125	15.5	10.5	15.7





#33

Benzo(k)fluoranthene

Concen: 0.21 ng

RT: 23.41 min Scan# 1484

Delta R.T. 0.00 min

Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.2

Tgt Ion:	252	Resp:	4961
Ion Ratio	Lower	Upper	
252	100		
253	24.6	20.3	30.5
125	14.3	9.6	14.4

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:07 PM



#34

Benzo(a)pyrene

Concen: 0.20 ng




RT: 23.98 min Scan# 1534

Delta R.T. 0.00 min

Lab File: BE092343.D

Acq: 2 Sep 2016 14:58

Tgt Ion:	252	Resp:	4487
Ion Ratio	Lower	Upper	
252	100		
253	26.4	19.0	28.6
125	15.6	11.8	17.8









#35
Dibenzo(a,h)anthracene
Concen: 0.18 ng
RT: 26.60 min Scan# 1715
Delta R.T. 0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Instrument :
BNA_E
ClientSampled :
SSTDICC0.2

Tgt Ion	Ratio	Lower	Upper
278	100		
139	22.3	13.8	20.8#
279	28.2	21.4	32.2

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:07 PM



#36
Benzo(g,h,i)perylene
Concen: 0.20 ng
RT: 27.35 min Scan# 1759
Delta R.T. 0.00 min
Lab File: BE092343.D
Acq: 2 Sep 2016 14:58

Tgt Ion	Ratio	Lower	Upper
276	100		
277	27.6	19.8	29.8
138	24.4	19.8	29.6

