

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090216\
 Data File : BE092347.D
 Acq On : 2 Sep 2016 17:24
 Operator : UM/SJ
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Manual Integrations
 APPROVED

sohil
 9/6/2016 1:15:18 PM

Quant Time: Sep 02 18:17:16 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 18:16:14 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	9981	5.00	ng	-0.02
7) Naphthalene-d8	10.95	136	49314	5.00	ng	0.00
12) Acenaphthene-d10	14.82	164	29465	5.00	ng	0.00
18) Phenanthrene-d10	17.58	188	67273	5.00	ng	0.02
24) Chrysene-d12	21.75	240	100183	5.00	ng	0.00
31) Perylene-d12	24.11	264	90849	5.00	ng	0.02

System Monitoring Compounds						
4) 2-Fluorophenol	5.68	112	7454	3.29	ng	-0.01
5) Phenol-d6	7.33	99	10798	3.64	ng	-0.02
8) Nitrobenzene-d5	9.34	82	10828	3.18	ng	-0.02
13) 2,4,6-Tribromophenol	16.31	330	3453	3.93	ng	-0.01
14) 2-Fluorobiphenyl	13.43	172	27539	3.01	ng	-0.01
26) Terphenyl-d14	20.17	244	41362	2.63	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.42	88	3761	2.46	ng	93
3) n-Nitrosodimethylamine	3.77	42	5051	3.98	ng	99
6) bis(2-Chloroethyl)ether	7.59	93	9674	3.90	ng	# 90
9) Naphthalene	10.99	128	32076	2.83	ng	# 95
10) Hexachlorobutadiene	11.30	225	4902	2.74	ng	100
11) 2-Methylnaphthalene	12.62	142	21762	3.01	ng	95
15) Acenaphthylene	14.53	152	157398	3.13	ng	100
16) Acenaphthene	14.88	154	24039	2.83	ng	97
17) Fluorene	15.87	166	31656	3.06	ng	100
19) Hexachlorobenzene	16.89	284	10094	3.13	ng	# 62
20) Pentachlorophenol	17.23	266	3566	5.61	ng	# 86
21) Phenanthrene	17.60	178	57465	2.95	ng	# 100
22) Anthracene	17.69	178	53955	3.37	ng	99
23) Fluoranthene	19.61	202	259471	3.17	ng	100
25) Pyrene	19.97	202	275858	2.70	ng	100
27) Benzo(a)anthracene	21.72	228	308241	2.74	ng	# 83
28) Chrysene	21.77	228	336318m	2.83	ng	
29) Bis(2-ethylhexyl)phthalate	21.63	149	46927	3.11	ng	# 75
30) Indeno(1,2,3-cd)pyrene	26.60	276	320933	2.65	ng	98
32) Benzo(b)fluoranthene	23.37	252	72132	3.08	ng	# 95
33) Benzo(k)fluoranthene	23.42	252	73112	2.96	ng	94
34) Benzo(a)pyrene	24.00	252	70848	3.10	ng	# 95
35) Dibenzo(a,h)anthracene	26.61	278	66708	3.20	ng	95
36) Benzo(g,h,i)perylene	27.36	276	271268	3.02	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090216\
Data File : BE092347.D
Acq On : 2 Sep 2016 17:24
Operator : UM/SJ
Sample : SSTDIC3.2
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleID :
SSTDIC3.2

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM

Quant Time: Sep 02 18:17:16 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 18:16:14 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.16 min Scan# 576

Delta R.T. -0.02 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
152	100		
150	190.1	106.0	159.0#
115	98.3	31.2	46.8#

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM



#2

1,4-Dioxane

Concen: 2.46 ng




RT: 3.42 min Scan# 72


Delta R.T. 0.00 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Tgt Ion	Ratio	Lower	Upper
88	100		
58	86.9	63.6	95.4
43	36.2	28.0	42.0





#3

n-Nitrosodimethylamine

Concen: 3.98 ng

RT: 3.77 min Scan# 102

Delta R.T. -0.03 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tgt Ion:	42	Resp:	5051
Ion Ratio	Lower	Upper	
42	100		
74	106.9	86.0	129.0
44	6.4	5.1	7.7

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM



#4

2-Fluorophenol

Concen: 3.29 ng




RT: 5.68 min Scan# 364


Delta R.T. -0.01 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Tgt Ion:	112	Resp:	7454
Ion Ratio	Lower	Upper	
112	100		
64	67.0	53.5	80.3
63	35.7	27.9	41.9





#5

Phenol-d6

Concen: 3.64 ng

RT: 7.33 min Scan# 540

Delta R.T. -0.02 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Instrument :

BNA_E


ClientSampled :

SSTDICC3.2




Tgt Ion:	99	Resp:	10798
Ion Ratio	Lower	Upper	
99	100		
42	23.2	0.0	0.0#
71	33.3	30.6	45.8

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 3.90 ng


RT: 7.59 min Scan# 551

Delta R.T. -0.02 min



Lab File: BE092347.D


Acq: 2 Sep 2016 17:24

Tgt Ion:	93	Resp:	9674
Ion Ratio	Lower	Upper	
93	100		
63	80.0	71.6	107.4
95	32.3	0.0	0.0#



Abundance Ion 93.00 (92.70 to 93.70): BE0
Ion 63.00 (62.70 to 63.70): BE0
Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.95 min Scan# 708

Delta R.T. 0.00 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Instrument :

BNA_E


ClientSampled :

SSTDICC3.2

Tgt Ion	Ratio	Resp	Lower	Upper
136	100	49314		
137	10.8	8.2	12.4	
54	7.7	9.6	14.4	#
68	5.9	6.7	10.1	#

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM




Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E


Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0

Time-->



Time-->



#8

Nitrobenzene-d5

Concen: 3.18 ng


RT: 9.34 min Scan# 627

Delta R.T. -0.02 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	10828		
128	32.4	39.0	58.4	#
54	47.8	44.8	67.2	#




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0

Time-->



Time-->



#9

Naphthalene

Concen: 2.83 ng

RT: 10.99 min Scan# 710

Delta R.T. -0.02 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tgt Ion:	128	Resp:	32076
Ion Ratio		Lower	Upper
128	100		
129	10.1	11.2	16.8#
127	15.0	11.6	17.4

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM



#10

Hexachlorobutadiene

Concen: 2.74 ng




RT: 11.30 min Scan# 726

Delta R.T. 0.00 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Tgt Ion:	225	Resp:	4902
Ion Ratio		Lower	Upper
225	100		
223	62.8	50.6	76.0
227	64.4	51.4	77.0









#11
2-Methylnaphthalene
Concen: 3.01 ng
RT: 12.62 min Scan# 814
Delta R.T. -0.02 min
Lab File: BE092347.D
Acq: 2 Sep 2016 17:24

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion:142 Resp: 21762
Ion Ratio Lower Upper
142 100
141 89.0 73.7 110.5
115 36.4 34.0 51.0




Manual Integrations
APPROVED


sohil
9/6/2016 1:15:18 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.82 min Scan# 980
Delta R.T. 0.00 min
Lab File: BE092347.D
Acq: 2 Sep 2016 17:24

Tgt Ion:164 Resp: 29465
Ion Ratio Lower Upper
164 100
162 108.6 71.4 107.0#
160 49.9 27.4 41.2#









#13
 2,4,6-Tribromophenol
 Concen: 3.93 ng
 RT: 16.31 min Scan# 1088
 Delta R.T. -0.01 min
 Lab File: BE092347.D
 Acq: 2 Sep 2016 17:24

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion: 330 Resp: 3453
 Ion Ratio Lower Upper
 330 100
 332 95.9 75.4 113.0
 141 41.3 31.0 46.6




Manual Integrations
 APPROVED


sohil
 9/6/2016 1:15:18 PM



#14
 2-Fluorobiphenyl
 Concen: 3.01 ng
 RT: 13.43 min Scan# 884
 Delta R.T. -0.01 min
 Lab File: BE092347.D
 Acq: 2 Sep 2016 17:24

Tgt Ion: 172 Resp: 27539
 Ion Ratio Lower Upper
 172 100
 171 36.9 30.1 45.1
 170 26.5 21.8 32.6









#15
Acenaphthylene
Concen: 3.13 ng
RT: 14.53 min Scan# 963
Delta R.T. 0.00 min
Lab File: BE092347.D
Acq: 2 Sep 2016 17:24

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
152	100		
151	5.0	3.9	5.9
153	13.0	10.4	15.6




Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM



#16
Acenaphthene
Concen: 2.83 ng
RT: 14.88 min Scan# 984
Delta R.T. 0.00 min
Lab File: BE092347.D
Acq: 2 Sep 2016 17:24

Tgt Ion	Ratio	Lower	Upper
154	100		
153	446.9	350.8	526.2
152	218.4	171.1	256.7









#17
Fluorene
 Concen: 3.06 ng
 RT: 15.87 min Scan# 1050
 Delta R.T. -0.00 min
 Lab File: BE092347.D
 Acq: 2 Sep 2016 17:24

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.9	78.2	117.4
167	13.4	11.0	16.4




Manual Integrations
 APPROVED

sohil
 9/6/2016 1:15:18 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.58 min Scan# 1154
 Delta R.T. 0.02 min
 Lab File: BE092347.D
 Acq: 2 Sep 2016 17:24

Tgt Ion	Ratio	Lower	Upper
188	100		
94	2.2	3.2	4.8#
80	1.7	2.6	3.8#









#19
Hexachlorobenzene
Concen: 3.13 ng
RT: 16.89 min Scan# 1124
Delta R.T. -0.00 min
Lab File: BE092347.D
Acq: 2 Sep 2016 17:24

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion	Ratio	Resp	Lower	Upper
284	100	10094		
142	0.0	29.1	43.7#	
249	27.2	27.9	41.9#	




Manual Integrations
APPROVED


sohil
9/6/2016 1:15:18 PM



#20
Pentachlorophenol
Concen: 5.61 ng
RT: 17.23 min Scan# 1139
Delta R.T. -0.02 min
Lab File: BE092347.D
Acq: 2 Sep 2016 17:24

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	3566		
268	61.4	62.5	93.7#	
264	64.3	57.2	85.8	





#21

Phenanthrene

Concen: 2.95 ng

RT: 17.60 min Scan# 1155

Delta R.T. -0.00 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tgt Ion:	178	Resp:	57465
Ion Ratio	Lower	Upper	
178	100		
176	19.8	15.8	23.8
179	14.9	0.0	0.0#

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM



#22

Anthracene

Concen: 3.37 ng




RT: 17.69 min Scan# 1159


Delta R.T. -0.00 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Tgt Ion:	178	Resp:	53955
Ion Ratio	Lower	Upper	
178	100		
176	19.6	15.0	22.4
179	15.1	12.2	18.2





#23

Fluoranthene

Concen: 3.17 ng

RT: 19.61 min Scan# 1266

Delta R.T. -0.00 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tgt	Ion	202	Resp:	259471
Ion	Ratio	Lower	Upper	
202	100			
101	2.9	2.2	3.4	
203	17.1	13.7	20.5	

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM



#24

Chrysene-d12

Concen: 5.00 ng




RT: 21.75 min Scan# 1375

Delta R.T. 0.00 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Tgt	Ion	240	Resp:	100183
Ion	Ratio	Lower	Upper	
240	100			
120	0.8	2.6	4.0#	
236	24.5	24.6	37.0#	









#25
 Pyrene
 Concen: 2.70 ng
 RT: 19.97 min Scan# 1287
 Delta R.T. -0.00 min
 Lab File: BE092347.D
 Acq: 2 Sep 2016 17:24

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	202	Resp:	275858
Ion Ratio	Lower	Upper	
202	100		
200	5.3	4.2	6.4
203	17.5	13.9	20.9




Manual Integrations
 APPROVED


sohil
 9/6/2016 1:15:18 PM



#26
 Terphenyl-d14
 Concen: 2.63 ng
 RT: 20.17 min Scan# 1299
 Delta R.T. -0.00 min
 Lab File: BE092347.D
 Acq: 2 Sep 2016 17:24

Tgt Ion:	244	Resp:	41362
Ion Ratio	Lower	Upper	
244	100		
212	8.8	6.7	10.1
122	5.8	3.6	5.4#





#27

Benzo(a)anthracene

Concen: 2.74 ng

RT: 21.72 min Scan# 1374

Delta R.T. 0.00 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Instrument :

BNA_E

ClientSampleId :

SSTDICC3.2

Tgt Ion:228 Resp: 308241
Ion Ratio Lower Upper

228 100


226 19.0 23.6 35.4#

229 22.5 13.3 19.9#




Manual Integrations
APPROVED

sohil

9/6/2016 1:15:18 PM



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E



#28

Chrysene

Concen: 2.83 ng m

RT: 21.77 min Scan# 1376

Delta R.T. 0.00 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24




Tgt Ion:228 Resp: 336318
Ion Ratio Lower Upper


228 100

226 31.2 36.3 54.5#

229 15.5 10.6 16.0

Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E









#29
 Bis(2-ethylhexyl)phthalate
 Concen: 3.11 ng
 RT: 21.63 min Scan# 1370
 Delta R.T. 0.00 min
 Lab File: BE092347.D
 Acq: 2 Sep 2016 17:24

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
149	100		
167	22.9	16.0	24.0
279	0.0	16.0	24.0




Manual Integrations
 APPROVED


sohil
 9/6/2016 1:15:18 PM



#30
 Indeno(1,2,3-cd)pyrene
 Concen: 2.65 ng
 RT: 26.60 min Scan# 1716
 Delta R.T. 0.02 min
 Lab File: BE092347.D
 Acq: 2 Sep 2016 17:24

Tgt Ion	Ratio	Lower	Upper
276	100		
138	26.7	20.3	30.5
277	24.9	19.7	29.5









#31
Perylene-d12
 Concen: 5.00 ng
 RT: 24.11 min Scan# 1546
 Delta R.T. 0.02 min
 Lab File: BE092347.D
 Acq: 2 Sep 2016 17:24

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.2	20.1	30.1
265	22.4	17.9	26.9




Manual Integrations
 APPROVED

sohil
 9/6/2016 1:15:18 PM



#32
Benzo(b)fluoranthene
 Concen: 3.08 ng
 RT: 23.37 min Scan# 1482
 Delta R.T. 0.01 min
 Lab File: BE092347.D
 Acq: 2 Sep 2016 17:24

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.9	18.7	28.1
125	10.2	10.5	15.7#





#33

Benzo(k)fluoranthene

Concen: 2.96 ng

RT: 23.42 min Scan# 1486

Delta R.T. 0.01 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24

Instrument :

BNA_E





ClientSampled :

SSTDICC3.2

Tgt Ion:	252	Resp:	73112
Ion Ratio	Lower	Upper	
252	100		
253	21.7	20.3	30.5
125	10.3	9.6	14.4

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM



#34

Benzo(a)pyrene

Concen: 3.10 ng




RT: 24.00 min Scan# 1536

Delta R.T. 0.01 min

Lab File: BE092347.D

Acq: 2 Sep 2016 17:24


Tgt Ion:	252	Resp:	70848
Ion Ratio	Lower	Upper	
252	100		
253	22.5	19.0	28.6
125	11.0	11.8	17.8#





#35
Dibenzo(a,h)anthracene
Concen: 3.20 ng
RT: 26.61 min Scan# 1717
Delta R.T. 0.02 min
Lab File: BE092347.D
Acq: 2 Sep 2016 17:24




Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2




Tgt Ion: 278 Resp: 66708
Ion Ratio Lower Upper
278 100
139 15.3 13.8 20.8
279 23.9 21.4 32.2

Manual Integrations
APPROVED

sohil
9/6/2016 1:15:18 PM



#36
Benzo(g,h,i)perylene
Concen: 3.02 ng
RT: 27.36 min Scan# 1761
Delta R.T. 0.02 min
Lab File: BE092347.D
Acq: 2 Sep 2016 17:24



Tgt Ion: 276 Resp: 271268
Ion Ratio Lower Upper
276 100
277 24.6 19.8 29.8
138 21.0 19.8 29.6

