

Data Path : Z:\HPCHEM1\BNA E\DATA\BE090415\  
 Data File : BE090737.D  
 Acq On : 4 Sep 2015 19:40  
 Operator : UM/IZ  
 Sample : G3562-04  
 Misc :  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 090215-D534-F-D-S

Quant Time: Sep 07 01:09:03 2015  
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090115.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Sep 02 11:40:50 2015  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc | Units | Dev(Min)       |
|--------------------------------|-------|------|----------|------|-------|----------------|
| 1) 1,4-Dichlorobenzene-d4      | 7.55  | 152  | 43771    | 5.00 | ng    | 0.00           |
| 6) Naphthalene-d8              | 10.31 | 136  | 188393   | 5.00 | ng    | 0.00           |
| 12) Acenaphthene-d10           | 14.17 | 164  | 102045   | 5.00 | ng    | 0.00           |
| 18) Phenanthrene-d10           | 16.91 | 188  | 256301   | 5.00 | ng    | 0.00           |
| 25) Chrysene-d12               | 21.07 | 240  | 328161   | 5.00 | ng    | 0.00           |
| 32) Perylene-d12               | 23.36 | 264  | 307646   | 5.00 | ng    | 0.00           |
| System Monitoring Compounds    |       |      |          |      |       |                |
| 4) 2-Fluorophenol              | 5.18  | 112  | 34366    | 4.04 | ng    | 0.00           |
| 5) Phenol-d6                   | 6.75  | 99   | 28052    | 2.30 | ng    | 0.00           |
| 7) Nitrobenzene-d5             | 8.71  | 82   | 47683    | 3.83 | ng    | 0.00           |
| 13) 2,4,6-Tribromophenol       | 15.67 | 330  | 27482    | 9.15 | ng    | 0.00           |
| 14) 2-Fluorobiphenyl           | 12.80 | 172  | 100162   | 3.54 | ng    | 0.00           |
| 27) Terphenyl-d14              | 19.53 | 244  | 195186   | 5.38 | ng    | -0.01          |
| Target Compounds               |       |      |          |      |       |                |
| 30) Bis(2-ethylhexyl)phthalate | 21.00 | 149  | 5682     | 0.36 | ng    | Qvalue<br># 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE090415\  
Data File : BE090737.D  
Acq On : 4 Sep 2015 19:40  
Operator : UM/IZ  
Sample : G3562-04  
Misc :  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
090215-D534-F-D-S

Quant Time: Sep 07 01:09:03 2015  
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090115.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Sep 02 11:40:50 2015  
Response via : Initial Calibration

