

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090415\
 Data File : BE090728.D
 Acq On : 4 Sep 2015 14:00
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 04 14:33:07 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.814	0.727	10.7	89	0.00
3	n-Nitrosodimethylamine	1.033	0.892	13.6	88	0.00
4 S	2-Fluorophenol	1.039	1.156	-11.3	115	0.00
5 S	Phenol-d6	1.392	1.616	-16.1	118	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
7 S	Nitrobenzene-d5	0.330	0.336	-1.8	108	0.00
8	Nitrobenzene	0.365	0.360	1.4	109	0.00
9	Naphthalene	1.091	1.093	-0.2	107	0.00
10	Hexachlorobutadiene	0.171	0.162	5.3	98	0.00
11	2-Methylnaphthalene	0.590	0.677	-14.7	126	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
13 S	2,4,6-Tribromophenol	0.143	0.173	-21.0	170#	0.00
14 S	2-Fluorobiphenyl	1.385	1.397	-0.9	121	0.00
15	Acenaphthylene	7.992	8.582	-7.4	132	0.00
16	Acenaphthene	1.335	1.353	-1.3	119	0.00
17	Fluorene	1.666	1.676	-0.6	120	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
19	4-Bromophenyl-phenylether	0.170	0.189	-11.2	128	0.00
20	Hexachlorobenzene	0.928	0.917	1.2	111	0.00
21	Pentachlorophenol	0.054	0.068	-25.9#	159#	0.00
22	Phenanthrene	1.237	1.305	-5.5	117	-0.02
23	Anthracene	1.035	1.159	-12.0	129	0.00
24	Fluoranthene	1.304	1.428	-9.5	131	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
26	Pyrene	1.035	1.242	-20.0	136	0.00
27 S	Terphenyl-d14	0.553	0.641	-15.9	130	0.00
28	Benzo(a)anthracene	1.156	1.213	-4.9	119	0.00
29	Chrysene	1.279	1.342	-4.9	114	0.00
30	Bis(2-ethylhexyl)phthalate	0.345	0.451	-30.7#	191#	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	1.351	-14.4	133	0.00
32 I	Perylene-d12	1.000	1.000	0.0	116	0.00
33	Benzo(b)fluoranthene	1.084	1.134	-4.6	124	0.00
34	Benzo(k)fluoranthene	1.263	1.375	-8.9	128	0.00
35 C	Benzo(a)pyrene	0.970	1.112	-14.6	142	0.00
36	Dibenzo(a,h)anthracene	1.020	1.124	-10.2	133	0.00
37	Benzo(g,h,i)perylene	1.116	1.186	-6.3	130	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0