

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090415\
 Data File : BE090728.D
 Acq On : 4 Sep 2015 14:00
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 04 14:33:07 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	97	0.00
2	1,4-Dioxane	1.000	0.974	2.6	89	0.00
3	n-Nitrosodimethylamine	1.000	0.940	6.0	88	0.00
4 S	2-Fluorophenol	1.000	1.201	-20.1	115	0.00
5 S	Phenol-d6	1.000	1.161	-16.1	118	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	106	0.00
7 S	Nitrobenzene-d5	1.000	1.016	-1.6	108	0.00
8	Nitrobenzene	1.000	0.977	2.3	109	0.00
9	Naphthalene	1.000	1.002	-0.2	107	0.00
10	Hexachlorobutadiene	1.000	0.952	4.8	98	0.00
11	2-Methylnaphthalene	1.000	1.148	-14.8	126	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	118	0.00
13 S	2,4,6-Tribromophenol	1.000	1.311	-31.1#	170	0.00
14 S	2-Fluorobiphenyl	1.000	1.009	-0.9	121	0.00
15	Acenaphthylene	1.000	1.074	-7.4	132	0.00
16	Acenaphthene	1.000	1.014	-1.4	119	0.00
17	Fluorene	1.000	1.006	-0.6	120	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	111	0.00
19	4-Bromophenyl-phenylether	1.000	1.107	-10.7	128	0.00
20	Hexachlorobenzene	1.000	1.028	-2.8	111	0.00
21	Pentachlorophenol	1.000	1.379	-37.9#	159	0.00
22	Phenanthrene	1.000	1.054	-5.4	117	-0.02
23	Anthracene	1.000	1.120	-12.0	129	0.00
24	Fluoranthene	1.000	1.095	-9.5	131	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	107	0.00
26	Pyrene	1.000	1.200	-20.0	136	0.00
27 S	Terphenyl-d14	1.000	1.160	-16.0	130	0.00
28	Benzo(a)anthracene	1.000	1.095	-9.5	119	0.00
29	Chrysene	1.000	1.105	-10.5	114	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.500	-50.0#	191	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.144	-14.4	133	0.00
32 I	Perylene-d12	5.000	5.000	0.0	116	0.00
33	Benzo(b)fluoranthene	1.000	1.046	-4.6	124	0.00
34	Benzo(k)fluoranthene	1.000	1.089	-8.9	128	0.00
35 C	Benzo(a)pyrene	1.000	1.147	-14.7	142	0.00
36	Dibenzo(a,h)anthracene	1.000	1.063	-6.3	133	0.00
37	Benzo(g,h,i)perylene	1.000	1.063	-6.3	130	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0