

Data Path : Z:\HPCHEM1\BNA E\DATA\BE090415\
 Data File : BE090730.D
 Acq On : 4 Sep 2015 15:23
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 04 16:04:21 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.814	0.759	6.8	79	0.00
3	n-Nitrosodimethylamine	1.033	0.853	17.4	72	0.00
4 S	2-Fluorophenol	1.039	0.999	3.8	84	0.00
5 S	Phenol-d6	1.392	1.440	-3.4	90	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
7 S	Nitrobenzene-d5	0.330	0.319	3.3	86	0.00
8	Nitrobenzene	0.365	0.348	4.7	88	0.00
9	Naphthalene	1.091	1.102	-1.0	90	0.00
10	Hexachlorobutadiene	0.171	0.167	2.3	84	0.00
11	2-Methylnaphthalene	0.590	0.665	-12.7	104	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
13 S	2,4,6-Tribromophenol	0.143	0.164	-14.7	137	0.00
14 S	2-Fluorobiphenyl	1.385	1.379	0.4	102	0.00
15	Acenaphthylene	7.992	8.448	-5.7	111	0.00
16	Acenaphthene	1.335	1.352	-1.3	101	0.00
17	Fluorene	1.666	1.680	-0.8	103	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
19	4-Bromophenyl-phenylether	0.170	0.173	-1.8	107	0.00
20	Hexachlorobenzene	0.928	0.890	4.1	99	0.00
21	Pentachlorophenol	0.054	0.054	0.0	116	0.00
22	Phenanthrene	1.237	1.259	-1.8	103	-0.02
23	Anthracene	1.035	1.093	-5.6	111	0.00
24	Fluoranthene	1.304	1.382	-6.0	116	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
26	Pyrene	1.035	1.187	-14.7	121	0.00
27 S	Terphenyl-d14	0.553	0.615	-11.2	116	0.00
28	Benzo(a)anthracene	1.156	1.190	-2.9	108	0.00
29	Chrysene	1.279	1.345	-5.2	106	0.00
30	Bis(2-ethylhexyl)phthalate	0.345	0.379	-9.9	149	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	1.349	-14.2	123	0.00
32 I	Perylene-d12	1.000	1.000	0.0	106	0.00
33	Benzo(b)fluoranthene	1.084	1.139	-5.1	114	0.00
34	Benzo(k)fluoranthene	1.263	1.357	-7.4	115	0.00
35 C	Benzo(a)pyrene	0.970	1.072	-10.5	125	0.00
36	Dibenzo(a,h)anthracene	1.020	1.135	-11.3	123	0.00
37	Benzo(g,h,i)perylene	1.116	1.184	-6.1	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0