

Data Path : Z:\HPCHEM1\BNA E\DATA\BE090415\
 Data File : BE090730.D
 Acq On : 4 Sep 2015 15:23
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 04 16:04:21 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	82	0.00
2	1,4-Dioxane	1.000	1.024	-2.4	79	0.00
3	n-Nitrosodimethylamine	1.000	0.895	10.5	72	0.00
4 S	2-Fluorophenol	1.000	1.041	-4.1	84	0.00
5 S	Phenol-d6	1.000	1.034	-3.4	90	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	89	0.00
7 S	Nitrobenzene-d5	1.000	0.966	3.4	86	0.00
8	Nitrobenzene	1.000	0.946	5.4	88	0.00
9	Naphthalene	1.000	1.010	-1.0	90	0.00
10	Hexachlorobutadiene	1.000	0.980	2.0	84	0.00
11	2-Methylnaphthalene	1.000	1.128	-12.8	104	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	101	0.00
13 S	2,4,6-Tribromophenol	1.000	1.255	-25.5#	137	0.00
14 S	2-Fluorobiphenyl	1.000	0.996	0.4	102	0.00
15	Acenaphthylene	1.000	1.057	-5.7	111	0.00
16	Acenaphthene	1.000	1.012	-1.2	101	0.00
17	Fluorene	1.000	1.009	-0.9	103	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	101	0.00
19	4-Bromophenyl-phenylether	1.000	1.014	-1.4	107	0.00
20	Hexachlorobenzene	1.000	0.994	0.6	99	0.00
21	Pentachlorophenol	1.000	1.193	-19.3	116	0.00
22	Phenanthrene	1.000	1.017	-1.7	103	-0.02
23	Anthracene	1.000	1.056	-5.6	111	0.00
24	Fluoranthene	1.000	1.060	-6.0	116	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	100	0.00
26	Pyrene	1.000	1.147	-14.7	121	0.00
27 S	Terphenyl-d14	1.000	1.111	-11.1	116	0.00
28	Benzo(a)anthracene	1.000	1.073	-7.3	108	0.00
29	Chrysene	1.000	1.107	-10.7	106	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.285	-28.5#	149	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.142	-14.2	123	0.00
32 I	Perylene-d12	5.000	5.000	0.0	106	0.00
33	Benzo(b)fluoranthene	1.000	1.051	-5.1	114	0.00
34	Benzo(k)fluoranthene	1.000	1.075	-7.5	115	0.00
35 C	Benzo(a)pyrene	1.000	1.105	-10.5	125	0.00
36	Dibenzo(a,h)anthracene	1.000	1.072	-7.2	123	0.00
37	Benzo(g,h,i)perylene	1.000	1.061	-6.1	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0