

Data Path : Z:\HPCHEM1\BNA E\DATA\BE090415\
 Data File : BE090745.D
 Acq On : 5 Sep 2015 00:29
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 07 01:03:25 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	1,4-Dioxane	0.814	0.768	5.7	76	0.00
3	n-Nitrosodimethylamine	1.033	0.888	14.0	71	0.00
4 S	2-Fluorophenol	1.039	1.059	-1.9	85	0.00
5 S	Phenol-d6	1.392	1.486	-6.8	88	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
7 S	Nitrobenzene-d5	0.330	0.324	1.8	83	0.00
8	Nitrobenzene	0.365	0.346	5.2	84	0.00
9	Naphthalene	1.091	1.096	-0.5	86	0.00
10	Hexachlorobutadiene	0.171	0.167	2.3	80	0.00
11	2-Methylnaphthalene	0.590	0.665	-12.7	99	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
13 S	2,4,6-Tribromophenol	0.143	0.148	-3.5	120	0.00
14 S	2-Fluorobiphenyl	1.385	1.353	2.3	97	0.00
15	Acenaphthylene	7.992	8.412	-5.3	107	0.00
16	Acenaphthene	1.335	1.367	-2.4	99	0.00
17	Fluorene	1.666	1.730	-3.8	103	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
19	4-Bromophenyl-phenylether	0.170	0.177	-4.1	103	0.00
20	Hexachlorobenzene	0.928	0.911	1.8	95	0.00
21	Pentachlorophenol	0.054	0.053	1.9	106	0.00
22	Phenanthrene	1.237	1.272	-2.8	98	-0.02
23	Anthracene	1.035	1.082	-4.5	103	0.00
24	Fluoranthene	1.304	1.378	-5.7	108	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
26	Pyrene	1.035	1.158	-11.9	113	0.00
27 S	Terphenyl-d14	0.553	0.596	-7.8	108	0.00
28	Benzo(a)anthracene	1.156	1.181	-2.2	104	0.00
29	Chrysene	1.279	1.335	-4.4	101	0.00
30	Bis(2-ethylhexyl)phthalate	0.345	0.335	2.9	127	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	1.337	-13.2	118	0.00
32 I	Perylene-d12	1.000	1.000	0.0	102	0.00
33	Benzo(b)fluoranthene	1.084	1.149	-6.0	111	0.00
34	Benzo(k)fluoranthene	1.263	1.310	-3.7	108	0.00
35 C	Benzo(a)pyrene	0.970	1.057	-9.0	119	0.00
36	Dibenzo(a,h)anthracene	1.020	1.110	-8.8	116	-0.01
37	Benzo(g,h,i)perylene	1.116	1.169	-4.7	113	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0