

Data Path : Z:\HPCHEM1\BNA E\DATA\BE090415\
 Data File : BE090745.D
 Acq On : 5 Sep 2015 00:29
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 07 01:03:25 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	78	0.00
2	1,4-Dioxane	1.000	1.038	-3.8	76	0.00
3	n-Nitrosodimethylamine	1.000	0.936	6.4	71	0.00
4 S	2-Fluorophenol	1.000	1.103	-10.3	85	0.00
5 S	Phenol-d6	1.000	1.067	-6.7	88	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	85	0.00
7 S	Nitrobenzene-d5	1.000	0.980	2.0	83	0.00
8	Nitrobenzene	1.000	0.941	5.9	84	0.00
9	Naphthalene	1.000	1.005	-0.5	86	0.00
10	Hexachlorobutadiene	1.000	0.979	2.1	80	0.00
11	2-Methylnaphthalene	1.000	1.128	-12.8	99	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	98	0.00
13 S	2,4,6-Tribromophenol	1.000	1.142	-14.2	120	0.00
14 S	2-Fluorobiphenyl	1.000	0.977	2.3	97	0.00
15	Acenaphthylene	1.000	1.053	-5.3	107	0.00
16	Acenaphthene	1.000	1.024	-2.4	99	0.00
17	Fluorene	1.000	1.039	-3.9	103	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	95	0.00
19	4-Bromophenyl-phenylether	1.000	1.039	-3.9	103	0.00
20	Hexachlorobenzene	1.000	1.021	-2.1	95	0.00
21	Pentachlorophenol	1.000	1.172	-17.2	106	0.00
22	Phenanthrene	1.000	1.028	-2.8	98	-0.02
23	Anthracene	1.000	1.046	-4.6	103	0.00
24	Fluoranthene	1.000	1.056	-5.6	108	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	96	0.00
26	Pyrene	1.000	1.119	-11.9	113	0.00
27 S	Terphenyl-d14	1.000	1.076	-7.6	108	0.00
28	Benzo(a)anthracene	1.000	1.066	-6.6	104	0.00
29	Chrysene	1.000	1.099	-9.9	101	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.154	-15.4	127	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.132	-13.2	118	0.00
32 I	Perylene-d12	5.000	5.000	0.0	102	0.00
33	Benzo(b)fluoranthene	1.000	1.059	-5.9	111	0.00
34	Benzo(k)fluoranthene	1.000	1.037	-3.7	108	0.00
35 C	Benzo(a)pyrene	1.000	1.089	-8.9	119	0.00
36	Dibenzo(a,h)anthracene	1.000	1.052	-5.2	116	-0.01
37	Benzo(g,h,i)perylene	1.000	1.048	-4.8	113	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0