

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE090517\
 Data File : BE093930.D
 Acq On : 5 Sep 2017 10:43
 Operator : SJ/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICCC0.4

Manual Integrations
 APPROVED

Sohil
 9/7/2017 6:30:43 PM

Quant Time: Sep 05 12:40:05 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 05 12:08:29 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	1879	0.40	ng	0.00
7) Naphthalene-d8	10.71	136	9676	0.40	ng	0.00
13) Acenaphthene-d10	14.52	164	5485	0.40	ng	0.00
19) Phenanthrene-d10	17.27	188	10518	0.40	ng	0.00
27) Chrysene-d12	21.43	240	13822	0.40	ng	0.00
34) Perylene-d12	23.94	264	11333	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.53	112	2400	0.41	ng	0.00
5) Phenol-d6	7.11	99	3546	0.40	ng	0.00
8) Nitrobenzene-d5	9.08	82	3259	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.30	152	6354	0.42	ng	0.00
14) 2,4,6-Tribromophenol	16.03	330	704	0.48	ng	0.00
15) 2-Fluorobiphenyl	13.16	172	8880	0.41	ng	0.00
25) Fluoranthene-d10	19.28	212	61232	0.50	ng	0.00
29) Terphenyl-d14	19.87	244	10174	0.41	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.43	88	1024	0.397	ng	96
3) n-Nitrosodimethylamine	3.75	42	1212	0.422	ng	# 97
6) bis(2-Chloroethyl)ether	7.34	93	2967	0.436	ng	87
9) Naphthalene	10.76	128	44969	0.408	ng	100
10) Hexachlorobutadiene	11.04	225	1622	0.417	ng	98
12) 2-Methylnaphthalene	12.37	142	7519	0.408	ng	99
16) Acenaphthylene	14.24	152	10666	0.393	ng	99
17) Acenaphthene	14.58	154	7516	0.411	ng	98
18) Fluorene	15.58	166	9397	0.397	ng	100
20) 4-Bromophenyl-phenylether	16.45	248	2719	0.624	ng	85
21) Hexachlorobenzene	16.58	284	2656	0.700	ng	# 100
22) Pentachlorophenol	16.94	266	667	0.405	ng	97
23) Phenanthrene	17.30	178	15418m	0.577	ng	
24) Anthracene	17.40	178	13040m	0.409	ng	
26) Fluoranthene	19.31	202	18087	0.494	ng	99
28) Pyrene	19.67	202	18958	0.408	ng	100
30) Benzo(a)anthracene	21.40	228	17119	0.417	ng	98
31) Chrysene	21.46	228	17830	0.391	ng	98
32) Bis(2-ethylhexyl)phthalate	21.30	149	39429	0.456	ng	100
33) Indeno(1,2,3-cd)pyrene	26.61	276	11819	0.380	ng	98
35) Benzo(b)fluoranthene	23.17	252	14487	0.400	ng	97
36) Benzo(k)fluoranthene	23.21	252	17641m	0.427	ng	
37) Benzo(a)pyrene	23.83	252	14646	0.408	ng	95
38) Dibenzo(a,h)anthracene	26.63	278	9791	0.386	ng	# 91
39) Benzo(g,h,i)perylene	27.42	276	10254	0.371	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE090517\
Data File : BE093930.D
Acq On : 5 Sep 2017 10:43
Operator : SJ/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM

Quant Time: Sep 05 12:40:05 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 05 12:08:29 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 0.400 ng

RT: 7.92 min Scan# 402

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Instrument :

BNA_E

ClientSampleId :

SSTDICCC0.4

Tot Ion:152 Resp: 1879

Ion Ratio Lower Upper

152 100

150 155.0 121.8 182.6


115 70.9 52.6 79.0

Manual Integrations

APPROVED

Sohil




9/7/2017 6:30:43 PM



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.397 ng

RT: 3.43 min Scan# 45

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43


Tgt Ion: 88 Resp: 1024

Ion Ratio Lower Upper

88 100

58 86.4 65.4 98.0



43 30.7 23.9 35.9



Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 0.422 ng

RT: 3.75 min Scan# 71

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	42	Resp:	1212
Ion	Ratio	Lower	Upper
42	100		
74	129.0	104.0	156.0
44	10.1	13.8	20.8#

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



#4

2-Fluorophenol

Concen: 0.415 ng




RT: 5.53 min Scan# 212




Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Tgt Ion:	112	Resp:	2400
Ion	Ratio	Lower	Upper
112	100		
64	72.5	57.7	86.5
63	141.9	116.0	174.0





#5

Phenol-d6

Concen: 0.401 ng

RT: 7.11 min Scan# 338

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Tot Ion:	99	Resp:	3546
Ion	Ratio	Lower	Upper
99	100		
42	21.3	17.1	25.7
71	34.7	26.6	39.8

Instrument :





BNA_E

ClientSampleId :

SSTDICCC0.4

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.436 ng


RT: 7.34 min Scan# 356


Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Tgt Ion:	93	Resp:	2967
Ion	Ratio	Lower	Upper
93	100		
63	324.7	283.4	425.2
95	30.4	26.9	40.3





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.71 min Scan# 577

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.2	13.8
54	40.7	21.9	32.9#
68	8.3	4.6	7.0#

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



#8

Nitrobenzene-d5

Concen: 0.405 ng




RT: 9.08 min Scan# 480


Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Tgt Ion	Ratio	Lower	Upper
82	100		
128	153.0	91.4	137.0#
54	204.9	197.8	296.8





#9

Naphthalene

Concen: 0.408 ng

RT: 10.76 min Scan# 580

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	128	Resp:	44969
Ion Ratio	Lower	Upper	
128	100		
129	2.9	2.5	3.7
127	3.4	2.6	4.0

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



#10

Hexachlorobutadiene

Concen: 0.417 ng




RT: 11.04 min Scan# 597

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Tgt Ion:	225	Resp:	1622
Ion Ratio	Lower	Upper	
225	100		
223	63.7	49.4	74.0
227	63.9	50.5	75.7





#11

2-Methylnaphthalene-d10

Concen: 0.415 ng

RT: 12.30 min Scan# 826

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Instrument :

BNA_E

ClientSampleId :

SSTDICCC0.4

Tot Ion:152 Resp: 6354

Ion Ratio Lower Upper





152 100

151 18.6 15.4 23.0

Manual Integrations
APPROVED

Sohil

9/7/2017 6:30:43 PM



#12

2-Methylnaphthalene

Concen: 0.408 ng

RT: 12.37 min Scan# 845

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43




Tgt Ion:142 Resp: 7519


Ion Ratio Lower Upper

142 100

141 90.5 71.4 107.2

115 37.3 29.4 44.0









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.52 min Scan# 1111
 Delta R.T. 0.00 min
 Lab File: BE093930.D
 Acq: 5 Sep 2017 10:43

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.4

Tot Ion	Ratio	Lower	Upper
164	100		
162	105.6	80.8	121.2
160	48.5	37.0	55.4




Manual Integrations
 APPROVED


Sohil
 9/7/2017 6:30:43 PM



#14
 2,4,6-Tribromophenol
 Concen: 0.479 ng
 RT: 16.03 min Scan# 1205
 Delta R.T. 0.00 min
 Lab File: BE093930.D
 Acq: 5 Sep 2017 10:43

Tgt Ion	Ratio	Lower	Upper
330	100		
332	104.4	76.9	115.3
141	0.0	0.0	0.0









#15
2-Fluorobiphenyl
Concen: 0.407 ng
RT: 13.16 min Scan# 1019
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
172	100		
171	34.9	26.9	40.3
170	23.0	17.3	25.9




Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



#16
Acenaphthylene
Concen: 0.393 ng
RT: 14.24 min Scan# 1092
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.7	23.5
153	13.0	10.6	16.0






#17
Acenaphthene
Concen: 0.411 ng
RT: 14.58 min Scan# 1115
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
154	100		
153	113.8	91.8	137.6
152	53.8	44.9	67.3

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18
Fluorene
Concen: 0.397 ng
RT: 15.58 min Scan# 1182
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43




Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.5	79.4	119.0
167	54.4	43.5	65.3


Abundance

Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.27 min Scan# 1243
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Instrument :
BNA_E
ClientSampled :
SSTDICCC0.4

Tot Ion	Ratio	Resp	Lower	Upper
188	100			
94	4.7	16.1	24.1#	
80	4.6	18.2	27.2#	




Manual Integrations
APPROVED


Sohil
9/7/2017 6:30:43 PM



#20
4-Bromophenyl-phenylether
Concen: 0.624 ng
RT: 16.45 min Scan# 1218
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Tgt Ion	Ratio	Resp	Lower	Upper
248	100			
250	118.3	111.8	167.8	
141	35.9	27.1	40.7	





#21

Hexachlorobenzene

Concen: 0.700 ng

RT: 16.58 min Scan# 1222

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	284	Resp:	2656
Ion Ratio		Lower	Upper
284	100		
142	0.0	0.0	0.0
249	22.3	0.0	0.0#

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



#22

Pentachlorophenol

Concen: 0.405 ng




RT: 16.94 min Scan# 1233


Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Tgt Ion:	266	Resp:	667
Ion Ratio		Lower	Upper
266	100		
264	69.7	53.5	80.3
268	73.4	60.0	90.0









#23
Phenanthrene
Concen: 0.577 ng m
RT: 17.30 min Scan# 1244
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
178	100		
176	21.0	0.0	0.0#
179	17.6	12.6	18.8




Manual Integrations
APPROVED


Sohil
9/7/2017 6:30:43 PM



#24
Anthracene
Concen: 0.409 ng m
RT: 17.40 min Scan# 1247
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Tgt Ion	Ratio	Lower	Upper
178	100		
176	0.0	16.0	24.0#
179	13.7	0.0	0.0#





#25

Fluoranthene-d10

Concen: 0.505 ng

RT: 19.28 min Scan# 1432

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	212	Resp:	61232
Ion Ratio	Lower	Upper	
212	100		
106	3.4	2.7	4.1
104	1.9	1.6	2.4

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



#26

Fluoranthene

Concen: 0.494 ng




RT: 19.31 min Scan# 1436

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Tgt Ion:	202	Resp:	18087
Ion Ratio	Lower	Upper	
202	100		
101	12.2	9.7	14.5
203	16.6	13.7	20.5









#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.43 min Scan# 1655
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
240	100		
120	9.0	8.7	13.1
236	26.8	21.3	31.9




Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



#28
Pyrene
Concen: 0.408 ng
RT: 19.67 min Scan# 1485
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.9	16.6	24.8
203	17.8	14.2	21.4





#29

Terphenyl-d14

Concen: 0.407 ng

RT: 19.87 min Scan# 1512

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	244	Resp:	10174
Ion Ratio	Lower	Upper	
244	100		
212	37.0	28.3	42.5
122	10.6	9.4	14.0

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



#30

Benzo(a)anthracene

Concen: 0.417 ng




RT: 21.40 min Scan# 1653


Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Tgt Ion:	228	Resp:	17119
Ion Ratio	Lower	Upper	
228	100		
226	27.1	20.8	31.2
229	20.1	16.6	25.0





#31

Chrysene

Concen: 0.391 ng

RT: 21.46 min Scan# 1658

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Instrument :

BNA_E


ClientSampled :

SSTDICCC0.4




Tot Ion:	228	Resp:	17830
Ion Ratio	Lower	Upper	
228	100		
226	30.0	24.6	37.0
229	21.1	15.8	23.8

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.456 ng

RT: 21.30 min Scan# 1644




Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Tgt Ion:	149	Resp:	39429
Ion Ratio	Lower	Upper	
149	100		
167	6.7	5.4	8.0
279	0.9	0.7	1.1

Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E









#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.380 ng
 RT: 26.61 min Scan# 2710
 Delta R.T. 0.00 min
 Lab File: BE093930.D
 Acq: 5 Sep 2017 10:43

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.4

Tot Ion	Ratio	Lower	Upper
276	100		
138	26.5	22.2	33.2
277	23.7	19.5	29.3




Manual Integrations
 APPROVED


Sohil
 9/7/2017 6:30:43 PM



#34
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.94 min Scan# 2128
 Delta R.T. 0.00 min
 Lab File: BE093930.D
 Acq: 5 Sep 2017 10:43

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.4	20.0	30.0
265	38.8	24.0	36.0#









#35
Benzo(b)fluoranthene
Concen: 0.400 ng
RT: 23.17 min Scan# 1958
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
252	100		
253	25.7	19.4	29.0
125	14.1	10.0	15.0




Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM



#36
Benzo(k)fluoranthene
Concen: 0.427 ng m
RT: 23.21 min Scan# 1968
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.0	18.1	27.1
125	14.8	9.8	14.6








#37
Benzo(a)pyrene
Concen: 0.408 ng
RT: 23.83 min Scan# 2103
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
252	100		
253	26.5	19.6	29.4
125	17.5	11.8	17.8





Manual Integrations
APPROVED




Sohil
9/7/2017 6:30:43 PM



#38
Dibenzo(a,h)anthracene
Concen: 0.386 ng
RT: 26.63 min Scan# 2714
Delta R.T. 0.00 min
Lab File: BE093930.D
Acq: 5 Sep 2017 10:43

Tgt Ion	Ratio	Lower	Upper
278	100		
139	27.3	16.7	25.1#
279	29.9	21.6	32.4





#39

Benzo(a,h,i)perylene

Concen: 0.371 ng

RT: 27.42 min Scan# 2889

Delta R.T. 0.00 min

Lab File: BE093930.D

Acq: 5 Sep 2017 10:43

Tot Ion:	276	Resp:	10254
Ion Ratio	Lower	Upper	
276	100		
277	27.0	20.6	30.8
138	28.1	21.9	32.9

Instrument :

BNA_E

ClientSampleId :

SSTDICCC0.4

Manual Integrations
APPROVED

Sohil
9/7/2017 6:30:43 PM

