

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE090617\
 Data File : BE093951.D
 Acq On : 6 Sep 2017 10:11
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 06 18:25:30 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 05 13:34:56 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	142	0.00
2	1,4-Dioxane	0.551	0.481	12.7	125	0.00
3	n-Nitrosodimethylamine	0.646	0.569	11.9	125	0.00
4 S	2-Fluorophenol	1.294	1.117	13.7	124	0.00
5 S	Phenol-d6	1.941	1.629	16.1	122	0.00
6	bis(2-Chloroethyl)ether	1.587	1.361	14.2	122	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	141	0.00
8 S	Nitrobenzene-d5	0.338	0.289	14.5	121	0.00
9	Naphthalene	4.585	3.930	14.3	119	0.00
10	Hexachlorobutadiene	0.167	0.143	14.4	120	0.00
11 SURR	2-Methylnaphthalene-d10	0.640	0.548	14.4	118	0.00
12	2-Methylnaphthalene	0.762	0.659	13.5	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.124	0.114	8.1	122	0.00
15 S	2-Fluorobiphenyl	1.582	1.384	12.5	117	0.00
16	Acenaphthylene	1.997	1.695	15.1	119	0.00
17	Acenaphthene	1.349	1.170	13.3	117	0.00
18	Fluorene	1.741	1.476	15.2	118	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	124	0.00
20	4-Bromophenyl-phenylether	0.219	0.228	-4.1	109	0.00
21	Hexachlorobenzene	0.236	0.217	8.1	107	0.00
22	Pentachlorophenol	0.063	0.067	-6.3	131	0.00
23	Phenanthrene	1.293	1.256	2.9	107	0.00
24	Anthracene	1.218	1.081	11.2	108	0.00
25 SURR	Fluoranthene-d10	5.960	5.312	10.9	113	0.00
26	Fluoranthene	1.783	1.560	12.5	113	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	133	-0.01
28	Pvrene	1.376	1.177	14.5	114	0.00
29 S	Terphenyl-d14	0.736	0.628	14.7	114	0.00
30	Benzo(a)anthracene	1.229	1.005	18.2	108	0.00
31	Chrysene	1.300	1.153	11.3	119	0.00
32	Bis(2-ethylhexyl)phthalate	2.983	2.407	19.3	113	0.00
33	Indeno(1,2,3-cd)pyrene	0.867	0.750	13.5	117	0.00
34 I	Perylene-d12	1.000	1.000	0.0	133	0.00
35	Benzo(b)fluoranthene	1.274	1.083	15.0	113	0.00
36	Benzo(k)fluoranthene	1.561	1.346	13.8	115	0.00
37 C	Benzo(a)pyrene	1.297	1.138	12.3	117	0.00
38	Dibenzo(a,h)anthracene	0.860	0.760	11.6	117	0.00
39	Benzo(a,h,i)perylene	0.907	0.787	13.2	116	0.00

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE090617\
Data File : BE093951.D
Acq On : 6 Sep 2017 10:11
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 06 18:25:30 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 05 13:34:56 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			