

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE090617\
 Data File : BE093951.D
 Acq On : 6 Sep 2017 10:11
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 06 18:25:30 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 05 13:34:56 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	142	0.00
2	1,4-Dioxane	0.400	0.349	12.8	125	0.00
3	n-Nitrosodimethylamine	0.400	0.353	11.8	125	0.00
4 S	2-Fluorophenol	0.400	0.345	13.8	124	0.00
5 S	Phenol-d6	0.400	0.336	16.0	122	0.00
6	bis(2-Chloroethyl)ether	0.400	0.343	14.2	122	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	141	0.00
8 S	Nitrobenzene-d5	0.400	0.342	14.5	121	0.00
9	Naphthalene	0.400	0.343	14.2	119	0.00
10	Hexachlorobutadiene	0.400	0.343	14.2	120	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.343	14.2	118	0.00
12	2-Methylnaphthalene	0.400	0.346	13.5	119	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.400	0.370	7.5	122	0.00
15 S	2-Fluorobiphenyl	0.400	0.350	12.5	117	0.00
16	Acenaphthylene	0.400	0.340	15.0	119	0.00
17	Acenaphthene	0.400	0.347	13.3	117	0.00
18	Fluorene	0.400	0.339	15.3	118	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	124	0.00
20	4-Bromophenyl-phenylether	0.400	0.390	2.5	109	0.00
21	Hexachlorobenzene	0.400	0.368	8.0	107	0.00
22	Pentachlorophenol	0.400	0.421	-5.2	131	0.00
23	Phenanthrene	0.400	0.361	9.8	107	0.00
24	Anthracene	0.400	0.355	11.3	108	0.00
25 SURR	Fluoranthene-d10	0.400	0.357	10.8	113	0.00
26	Fluoranthene	0.400	0.350	12.5	113	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	133	-0.01
28	Pvrene	0.400	0.342	14.5	114	0.00
29 S	Terphenyl-d14	0.400	0.342	14.5	114	0.00
30	Benzo(a)anthracene	0.400	0.327	18.3	108	0.00
31	Chrysene	0.400	0.355	11.3	119	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.323	19.3	113	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.346	13.5	117	0.00
34 I	Perylene-d12	0.400	0.400	0.0	133	0.00
35	Benzo(b)fluoranthene	0.400	0.340	15.0	113	0.00
36	Benzo(k)fluoranthene	0.400	0.345	13.8	115	0.00
37 C	Benzo(a)pyrene	0.400	0.351	12.3	117	0.00
38	Dibenzo(a,h)anthracene	0.400	0.354	11.5	117	0.00
39	Benzo(a,h,i)perylene	0.400	0.347	13.3	116	0.00

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE090617\
Data File : BE093951.D
Acq On : 6 Sep 2017 10:11
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 06 18:25:30 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 05 13:34:56 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				