

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090916\
 Data File : BE092356.D
 Acq On : 9 Sep 2016 15:57
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 09 17:24:14 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	160#	-0.02
2	1,4-Dioxane	0.707	0.648	8.3	142	0.00
3	n-Nitrosodimethylamine	0.661	0.599	9.4	173#	0.00
4 S	2-Fluorophenol	1.189	0.922	22.5	146	0.00
5 S	Phenol-d6	1.586	1.304	17.8	171#	0.00
6	bis(2-Chloroethyl)ether	1.288	1.078	16.3	169#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	161#	0.00
8 S	Nitrobenzene-d5	0.359	0.341	5.0	194#	0.00
9	Naphthalene	1.120	1.052	6.1	159#	0.00
10	Hexachlorobutadiene	0.175	0.168	4.0	157#	0.00
11	2-Methylnaphthalene	0.733	0.665	9.3	158#	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	156#	0.00
13 S	2,4,6-Tribromophenol	0.155	0.126	18.7	158#	0.00
14 S	2-Fluorobiphenyl	1.534	1.438	6.3	162#	0.00
15	Acenaphthylene	8.606	7.681	10.7	155#	0.00
16	Acenaphthene	1.393	1.280	8.1	156#	0.00
17	Fluorene	1.749	1.549	11.4	153#	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	155#	0.00
19	Hexachlorobenzene	0.239	0.210	12.1	150	0.00
20	Pentachlorophenol	0.053	0.034	35.8#	150	0.00
21	Phenanthrene	1.405	1.302	7.3	158#	0.00
22	Anthracene	1.250	1.063	15.0	159#	0.00
23	Fluoranthene	6.135	5.334	13.1	153#	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	161#	0.00
25	Pyrene	4.864	4.637	4.7	161#	0.00
26 S	Terphenyl-d14	0.740	0.718	3.0	166#	0.00
27	Benzo(a)anthracene	5.396	4.601	14.7	140	0.00
28	Chrysene	5.570	5.628	-1.0	180#	0.00
29	Bis(2-ethylhexyl)phthalate	0.749	0.561	25.1#	146	0.00
30	Indeno(1,2,3-cd)pyrene	5.684	5.021	11.7	145	0.02
31 I	Perylene-d12	1.000	1.000	0.0	149	0.01
32	Benzo(b)fluoranthene	1.284	1.130	12.0	148	0.01
33	Benzo(k)fluoranthene	1.357	1.232	9.2	140	0.01
34 C	Benzo(a)pyrene	1.265	1.122	11.3	148	0.01
35	Dibenzo(a,h)anthracene	1.170	1.003	14.3	144	0.02
36	Benzo(g,h,i)perylene	4.933	4.319	12.4	140	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0