

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090916\
 Data File : BE092367.D
 Acq On : 9 Sep 2016 22:42
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 10 03:26:55 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	169#	-0.02
2	1,4-Dioxane	0.707	0.634	10.3	147	0.00
3	n-Nitrosodimethylamine	0.661	0.633	4.2	193#	-0.01
4 S	2-Fluorophenol	1.189	1.135	4.5	190#	0.00
5 S	Phenol-d6	1.586	1.480	6.7	205#	0.00
6	bis(2-Chloroethyl)ether	1.288	1.251	2.9	208#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	181#	0.00
8 S	Nitrobenzene-d5	0.359	0.312	13.1	199#	-0.02
9	Naphthalene	1.120	1.060	5.4	179#	0.00
10	Hexachlorobutadiene	0.175	0.164	6.3	172#	0.00
11	2-Methylnaphthalene	0.733	0.708	3.4	188#	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	185#	0.00
13 S	2,4,6-Tribromophenol	0.155	0.154	0.6	230#	0.00
14 S	2-Fluorobiphenyl	1.534	1.433	6.6	192#	0.00
15	Acenaphthylene	8.606	7.907	8.1	190#	0.00
16	Acenaphthene	1.393	1.301	6.6	188#	0.00
17	Fluorene	1.749	1.615	7.7	190#	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	175#	0.02
19	Hexachlorobenzene	0.239	0.231	3.3	185#	0.00
20	Pentachlorophenol	0.053	0.050	5.7	243#	-0.02
21	Phenanthrene	1.405	1.283	8.7	176#	0.00
22	Anthracene	1.250	1.156	7.5	195#	0.00
23	Fluoranthene	6.135	5.578	9.1	180#	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	195#	0.00
25	Pyrene	4.864	4.386	9.8	185#	0.00
26 S	Terphenyl-d14	0.740	0.653	11.8	183#	0.00
27	Benzo(a)anthracene	5.396	5.025	6.9	186#	0.00
28	Chrysene	5.570	4.738	14.9	184#	0.00
29	Bis(2-ethylhexyl)phthalate	0.749	0.571	23.8	180#	0.00
30	Indeno(1,2,3-cd)pyrene	5.684	4.594	19.2	161#	0.02
31 I	Perylene-d12	1.000	1.000	0.0	157#	0.02
32	Benzo(b)fluoranthene	1.284	1.251	2.6	173#	0.02
33	Benzo(k)fluoranthene	1.357	1.219	10.2	146	0.02
34 C	Benzo(a)pyrene	1.265	1.186	6.2	165#	0.01
35	Dibenzo(a,h)anthracene	1.170	1.092	6.7	165#	0.03
36	Benzo(g,h,i)perylene	4.933	4.623	6.3	158#	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0