

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090916\
 Data File : BE092367.D
 Acq On : 9 Sep 2016 22:42
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

sohil
 9/12/2016 1:42:08 PM

Quant Time: Sep 10 03:26:55 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	12033	5.00	ng	-0.02
7) Naphthalene-d8	10.95	136	60614	5.00	ng	0.00
12) Acenaphthene-d10	14.82	164	39969	5.00	ng	0.00
18) Phenanthrene-d10	17.58	188	90983	5.00	ng	0.02
24) Chrysene-d12	21.75	240	126270	5.00	ng	0.00
31) Perylene-d12	24.11	264	109412	5.00	ng	0.02

System Monitoring Compounds

4) 2-Fluorophenol	5.68	112	1093	0.38	ng	0.00
5) Phenol-d6	7.36	99	1425	0.37	ng	0.00
8) Nitrobenzene-d5	9.34	82	1513	0.35	ng	-0.02
13) 2,4,6-Tribromophenol	16.32	330	494	0.40	ng	0.00
14) 2-Fluorobiphenyl	13.44	172	4583	0.37	ng	0.00
26) Terphenyl-d14	20.17	244	6600	0.35	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.42	88	610	0.33	ng	95
3) n-Nitrosodimethylamine	3.79	42	609	0.39	ng	# 93
6) bis(2-Chloroethyl)ether	7.59	93	1204	0.40	ng	97
9) Naphthalene	11.01	128	5141	0.38	ng	98
10) Hexachlorobutadiene	11.30	225	793	0.37	ng	99
11) 2-Methylnaphthalene	12.64	142	3434	0.39	ng	92
15) Acenaphthylene	14.53	152	25282	0.37	ng	100
16) Acenaphthene	14.88	154	4160	0.37	ng	97
17) Fluorene	15.87	166	5164	0.37	ng	99
19) Hexachlorobenzene	16.89	284	1678	0.39	ng	# 63
20) Pentachlorophenol	17.23	266	362	0.44	ng	89
21) Phenanthrene	17.60	178	9339	0.37	ng	# 95
22) Anthracene	17.70	178	8414	0.37	ng	99
23) Fluoranthene	19.61	202	40601	0.36	ng	98
25) Pyrene	19.97	202	44308	0.36	ng	100
27) Benzo(a)anthracene	21.72	228	50765m	0.37	ng	
28) Chrysene	21.77	228	47862m	0.27	ng	
29) Bis(2-ethylhexyl)phthalate	21.63	149	5764	0.31	ng	# 75
30) Indeno(1,2,3-cd)pyrene	26.60	276	46403	0.32	ng	98
32) Benzo(b)fluoranthene	23.38	252	10951	0.39	ng	98
33) Benzo(k)fluoranthene	23.43	252	10674	0.36	ng	97
34) Benzo(a)pyrene	23.99	252	10383	0.38	ng	98
35) Dibenzo(a,h)anthracene	26.63	278	9560	0.37	ng	98
36) Benzo(g,h,i)perylene	27.37	276	40462	0.37	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE090916\
Data File : BE092367.D
Acq On : 9 Sep 2016 22:42
Operator : UM/SJ
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

sohil
9/12/2016 1:42:08 PM

Quant Time: Sep 10 03:26:55 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 19:34:51 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.17 min Scan# 576

Delta R.T. -0.02 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
152	100		
150	186.1	106.0	159.0#
115	93.5	31.2	46.8#

Manual Integrations
APPROVED

sohil
9/12/2016 1:42:08 PM



#2

1,4-Dioxane

Concen: 0.33 ng




RT: 3.42 min Scan# 72


Delta R.T. -0.00 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
88	100		
58	83.6	63.6	95.4
43	38.9	28.0	42.0





#3

n-Nitrosodimethylamine

Concen: 0.39 ng

RT: 3.79 min Scan# 104

Delta R.T. -0.01 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	42	Resp:	609
Ion	Ratio	Lower	Upper
42	100		
74	113.1	86.0	129.0
44	18.7	5.1	7.7#

Manual Integrations
APPROVED

sohil
9/12/2016 1:42:08 PM



#4

2-Fluorophenol

Concen: 0.38 ng




RT: 5.68 min Scan# 365




Delta R.T. -0.01 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Tgt Ion:	112	Resp:	1093
Ion	Ratio	Lower	Upper
112	100		
64	67.3	53.5	80.3
63	36.0	27.9	41.9





#5

Phenol-d6

Concen: 0.37 ng

RT: 7.36 min Scan# 541

Delta R.T. 0.00 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
99	100		
42	0.0	16.0	24.0#
71	0.0	30.6	45.8#

Instrument :





BNA_E

ClientSampleId :

SSTDCCC0.4EC

Manual Integrations
APPROVED

sohil
9/12/2016 1:42:08 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.40 ng


RT: 7.59 min Scan# 551


Delta R.T. -0.02 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
93	100		
63	87.7	71.6	107.4
95	32.7	24.0	36.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.95 min Scan# 708

Delta R.T. -0.00 min

Lab File: BE092367.D


Acq: 9 Sep 2016 22:42

Instrument :

BNA_E

ClientSampled :




SSTDCCC0.4EC



Tgt Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.2	12.4
54	9.5	9.6	14.4#
68	6.9	6.7	10.1

Manual Integrations
APPROVED

sohil
9/12/2016 1:42:08 PM



#8

Nitrobenzene-d5

Concen: 0.35 ng




RT: 9.34 min Scan# 627


Delta R.T. -0.02 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
82	100		
128	23.0	39.0	58.4#
54	81.1	44.8	67.2#





#9

Naphthalene

Concen: 0.38 ng

RT: 11.01 min Scan# 711

Delta R.T. -0.00 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	128	Resp:	5141
Ion Ratio	Lower	Upper	
128	100		
129	13.4	11.2	16.8
127	13.6	11.6	17.4

Manual Integrations
APPROVED

sohil
9/12/2016 1:42:08 PM



#10

Hexachlorobutadiene

Concen: 0.37 ng




RT: 11.30 min Scan# 726


Delta R.T. -0.00 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Tgt Ion:	225	Resp:	793
Ion Ratio	Lower	Upper	
225	100		
223	63.6	50.6	76.0
227	65.2	51.4	77.0





#11

2-Methylnaphthalene

Concen: 0.39 ng

RT: 12.64 min Scan# 815

Delta R.T. -0.01 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	142	Resp:	3434
Ion Ratio	Lower	Upper	
142	100		
141	85.9	73.7	110.5
115	35.7	34.0	51.0

Manual Integrations
APPROVED

sohil
9/12/2016 1:42:08 PM



#12

Acenaphthene-d10

Concen: 5.00 ng




RT: 14.82 min Scan# 980


Delta R.T. -0.00 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Tgt Ion:	164	Resp:	39969
Ion Ratio	Lower	Upper	
164	100		
162	110.2	71.4	107.0#
160	52.5	27.4	41.2#









#13
 2,4,6-Tribromophenol
 Concen: 0.40 ng
 RT: 16.32 min Scan# 1089
 Delta R.T. 0.00 min
 Lab File: BE092367.D
 Acq: 9 Sep 2016 22:42

Instrument :
 BNA_E
 ClientSampleId :
 SSTCCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.4	75.4	113.0
141	41.7	31.0	46.6




Manual Integrations
 APPROVED


sohil
 9/12/2016 1:42:08 PM



#14
 2-Fluorobiphenyl
 Concen: 0.37 ng
 RT: 13.44 min Scan# 885
 Delta R.T. -0.00 min
 Lab File: BE092367.D
 Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
172	100		
171	33.7	30.1	45.1
170	21.7	21.8	32.6#









#15
Acenaphthylene
Concen: 0.37 ng
RT: 14.53 min Scan# 963
Delta R.T. -0.00 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.9	5.9
153	13.0	10.4	15.6




Manual Integrations
APPROVED


sohil
9/12/2016 1:42:08 PM



#16
Acenaphthene
Concen: 0.37 ng
RT: 14.88 min Scan# 984
Delta R.T. -0.00 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
154	100		
153	430.4	350.8	526.2
152	211.6	171.1	256.7









#17
Fluorene
 Concen: 0.37 ng
 RT: 15.87 min Scan# 1050
 Delta R.T. 0.00 min
 Lab File: BE092367.D
 Acq: 9 Sep 2016 22:42

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.3	78.2	117.4
167	13.4	11.0	16.4




Manual Integrations
 APPROVED


sohil
 9/12/2016 1:42:08 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.58 min Scan# 1154
 Delta R.T. 0.02 min
 Lab File: BE092367.D
 Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
188	100		
94	2.4	3.2	4.8#
80	2.0	2.6	3.8#






#19
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.89 min Scan# 1124
Delta R.T. 0.00 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion	Ratio	Lower	Upper
284	100		
142	0.0	29.1	43.7#
249	28.8	27.9	41.9

Manual Integrations
APPROVED

sohil
9/12/2016 1:42:08 PM




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#20
Pentachlorophenol
Concen: 0.44 ng
RT: 17.23 min Scan# 1139
Delta R.T. -0.02 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
266	100		
268	69.3	62.5	93.7
264	81.3	57.2	85.8



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E









#21
Phenanthrene
Concen: 0.37 ng
RT: 17.60 min Scan# 1155
Delta R.T. 0.00 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.9	15.8	23.8
179	15.5	16.0	24.0




Manual Integrations
APPROVED


sohil
9/12/2016 1:42:08 PM



#22
Anthracene
Concen: 0.37 ng
RT: 17.70 min Scan# 1159
Delta R.T. 0.00 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.0	22.4
179	14.8	12.2	18.2





#23

Fluoranthene

Concen: 0.36 ng

RT: 19.61 min Scan# 1266

Delta R.T. 0.00 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
202	100		
101	2.9	2.2	3.4
203	17.9	13.7	20.5

Manual Integrations
APPROVED

sohil
9/12/2016 1:42:08 PM



#24

Chrysene-d12

Concen: 5.00 ng




RT: 21.75 min Scan# 1375


Delta R.T. 0.00 min

Lab File: BE092367.D

Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
240	100		
120	1.5	2.6	4.0#
236	26.5	24.6	37.0









#25
 Pyrene
 Concen: 0.36 ng
 RT: 19.97 min Scan# 1287
 Delta R.T. 0.00 min
 Lab File: BE092367.D
 Acq: 9 Sep 2016 22:42

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	44308		
200	5.2		4.2	6.4
203	17.6		13.9	20.9




Manual Integrations
 APPROVED


sohil
 9/12/2016 1:42:08 PM



#26
 Terphenyl-d14
 Concen: 0.35 ng
 RT: 20.17 min Scan# 1299
 Delta R.T. 0.00 min
 Lab File: BE092367.D
 Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Resp	Lower	Upper
244	100	6600		
212	10.8		6.7	10.1#
122	14.5		3.6	5.4#









#27
Benzo(a)anthracene
Concen: 0.37 ng m
RT: 21.72 min Scan# 1374
Delta R.T. 0.00 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
228	100		
226	23.9	23.6	35.4
229	20.3	13.3	19.9




Manual Integrations
APPROVED


sohil
9/12/2016 1:42:08 PM



#28
Chrysene
Concen: 0.27 ng m
RT: 21.77 min Scan# 1376
Delta R.T. 0.00 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
228	100		
226	38.4	36.3	54.5
229	14.2	10.6	16.0









#29
 Bis(2-ethylhexyl)phthalate
 Concen: 0.31 ng
 RT: 21.63 min Scan# 1370
 Delta R.T. 0.00 min
 Lab File: BE092367.D
 Acq: 9 Sep 2016 22:42

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
149	100		
167	23.3	16.0	24.0
279	0.0	16.0	24.0




Manual Integrations
 APPROVED


sohil
 9/12/2016 1:42:08 PM



#30
 Indeno(1,2,3-cd)pyrene
 Concen: 0.32 ng
 RT: 26.60 min Scan# 1716
 Delta R.T. 0.02 min
 Lab File: BE092367.D
 Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
276	100		
138	27.4	20.3	30.5
277	24.9	19.7	29.5









#31
Perylene-d12
 Concen: 5.00 ng
 RT: 24.11 min Scan# 1546
 Delta R.T. 0.02 min
 Lab File: BE092367.D
 Acq: 9 Sep 2016 22:42

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Resp	Lower	Upper
264	100	109412		
260	25.7	20.1	30.1	
265	21.2	17.9	26.9	




Manual Integrations
 APPROVED


sohil
 9/12/2016 1:42:08 PM



#32
Benzo(b)fluoranthene
 Concen: 0.39 ng
 RT: 23.38 min Scan# 1483
 Delta R.T. 0.02 min
 Lab File: BE092367.D
 Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	10951		
253	24.0	18.7	28.1	
125	11.9	10.5	15.7	









#33
Benzo(k)fluoranthene
Concen: 0.36 ng
RT: 23.43 min Scan# 1487
Delta R.T. 0.02 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.2	20.3	30.5
125	12.0	9.6	14.4




Manual Integrations
APPROVED


sohil
9/12/2016 1:42:08 PM



#34
Benzo(a)pyrene
Concen: 0.38 ng
RT: 23.99 min Scan# 1536
Delta R.T. 0.01 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.0	19.0	28.6
125	16.3	11.8	17.8









#35
Dibenzo(a,h)anthracene
Concen: 0.37 ng
RT: 26.63 min Scan# 1718
Delta R.T. 0.03 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.1	13.8	20.8
279	26.0	21.4	32.2

Manual Integrations
APPROVED

sohil
9/12/2016 1:42:08 PM



#36
Benzo(g,h,i)perylene
Concen: 0.37 ng
RT: 27.37 min Scan# 1761
Delta R.T. 0.02 min
Lab File: BE092367.D
Acq: 9 Sep 2016 22:42

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.8	29.8
138	25.7	19.8	29.6

