

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091115\
 Data File : BE090783.D
 Acq On : 11 Sep 2015 11:55
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 11 23:56:57 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.814	0.611	24.9	90	0.00
3	n-Nitrosodimethylamine	1.033	0.758	26.6#	90	0.00
4 S	2-Fluorophenol	1.039	0.741	28.7#	88	0.00
5 S	Phenol-d6	1.392	1.163	16.5	102	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	129	0.00
7 S	Nitrobenzene-d5	0.330	0.267	19.1	104	0.00
8	Nitrobenzene	0.365	0.301	17.5	110	0.00
9	Naphthalene	1.091	0.936	14.2	111	-0.01
10	Hexachlorobutadiene	0.171	0.145	15.2	105	0.00
11	2-Methylnaphthalene	0.590	0.526	10.8	118	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	141	0.00
13 S	2,4,6-Tribromophenol	0.143	0.091	36.4#	106	0.00
14 S	2-Fluorobiphenyl	1.385	1.138	17.8	117	0.00
15	Acenaphthylene	7.992	6.649	16.8	122	0.00
16	Acenaphthene	1.335	1.143	14.4	119	0.00
17	Fluorene	1.666	1.360	18.4	116	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	133	0.00
19	4-Bromophenyl-phenylether	0.170	0.140	17.6	114	0.00
20	Hexachlorobenzene	0.928	0.768	17.2	111	0.00
21	Pentachlorophenol	0.054	0.030	44.4#	83	0.00
22	Phenanthrene	1.237	1.078	12.9	116	-0.02
23	Anthracene	1.035	0.878	15.2	117	0.00
24	Fluoranthene	1.304	1.018	21.9	112	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
26	Pyrene	1.035	0.929	10.2	118	0.00
27 S	Terphenyl-d14	0.553	0.479	13.4	112	0.00
28	Benzo(a)anthracene	1.156	0.972	15.9	110	0.00
29	Chrysene	1.279	1.117	12.7	110	0.00
30	Bis(2-ethylhexyl)phthalate	0.345	0.197	42.9#	96	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	1.024	13.3	116	0.00
32 I	Perylene-d12	1.000	1.000	0.0	128	0.00
33	Benzo(b)fluoranthene	1.084	0.905	16.5	110	0.00
34	Benzo(k)fluoranthene	1.263	1.087	13.9	112	0.00
35 C	Benzo(a)pyrene	0.970	0.817	15.8	115	0.00
36	Dibenzo(a,h)anthracene	1.020	0.878	13.9	115	-0.01
37	Benzo(g,h,i)perylene	1.116	0.900	19.4	109	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0