

Data Path : Z:\HPCHEM1\BNA_E\Data\BE091117\
 Data File : BE093954.D
 Acq On : 11 Sep 2017 9:43
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 12 07:52:20 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 05 13:34:56 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.551	0.562	-2.0	106	0.00
3	n-Nitrosodimethylamine	0.646	0.626	3.1	100	0.00
4 S	2-Fluorophenol	1.294	1.278	1.2	103	0.00
5 S	Phenol-d6	1.941	1.821	6.2	99	0.00
6	bis(2-Chloroethyl)ether	1.587	1.563	1.5	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.338	0.324	4.1	99	0.00
9	Naphthalene	4.585	4.645	-1.3	103	0.00
10	Hexachlorobutadiene	0.167	0.162	3.0	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.640	0.635	0.8	100	0.00
12	2-Methylnaphthalene	0.762	0.758	0.5	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.124	0.121	2.4	94	0.00
15 S	2-Fluorobiphenyl	1.582	1.633	-3.2	101	0.00
16	Acenaphthylene	1.997	1.904	4.7	98	0.00
17	Acenaphthene	1.349	1.368	-1.4	100	0.00
18	Fluorene	1.741	1.775	-2.0	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4-Bromophenyl-phenylether	0.219	0.260	-18.7	101	0.00
21	Hexachlorobenzene	0.236	0.229	3.0	91	0.00
22	Pentachlorophenol	0.063	0.064	-1.6	101	0.00
23	Phenanthrene	1.293	1.452	-12.3	99	0.00
24	Anthracene	1.218	1.233	-1.2	99	0.00
25 SURR	Fluoranthene-d10	5.960	5.814	2.4	100	0.00
26	Fluoranthene	1.783	1.724	3.3	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pyrene	1.376	1.407	-2.3	100	0.00
29 S	Terphenyl-d14	0.736	0.744	-1.1	99	0.00
30	Benzo(a)anthracene	1.229	1.172	4.6	93	0.00
31	Chrysene	1.300	1.309	-0.7	99	0.00
32	Bis(2-ethylhexyl)phthalate	2.983	2.811	5.8	96	0.00
33	Indeno(1,2,3-cd)pyrene	0.867	0.703	18.9	80	0.00
34 I	Perylene-d12	1.000	1.000	0.0	90	0.00
35	Benzo(b)fluoranthene	1.274	1.209	5.1	85	0.00
36	Benzo(k)fluoranthene	1.561	1.631	-4.5	94	0.00
37 C	Benzo(a)pyrene	1.297	1.276	1.6	89	0.00
38	Dibenzo(a,h)anthracene	0.860	0.731	15.0	76	0.00
39	Benzo(g,h,i)perylene	0.907	0.808	10.9	80	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE091117\
Data File : BE093954.D
Acq On : 11 Sep 2017 9:43
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 12 07:52:20 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 05 13:34:56 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		