

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091216\
 Data File : BE092378.D
 Acq On : 12 Sep 2016 19:38
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 13 05:32:55 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	150	-0.02
2	1,4-Dioxane	0.400	0.326	18.5	131	0.00
3	n-Nitrosodimethylamine	0.400	0.494	-23.5	228	-0.03
4 S	2-Fluorophenol	0.400	0.478	-19.5	211	-0.01
5 S	Phenol-d6	0.400	0.525	-31.3#	256	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.463	-15.8	222	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	157	0.00
8 S	Nitrobenzene-d5	0.400	0.493	-23.2	245	-0.02
9	Naphthalene	0.400	0.402	-0.5	165	-0.02
10	Hexachlorobutadiene	0.400	0.410	-2.5	163	-0.02
11	2-Methylnaphthalene	0.400	0.408	-2.0	173	-0.02
12 I	Acenaphthene-d10	5.000	5.000	0.0	148	0.00
13 S	2,4,6-Tribromophenol	0.400	0.532	-33.0#	245	-0.01
14 S	2-Fluorobiphenyl	0.400	0.392	2.0	161	-0.01
15	Acenaphthylene	0.400	0.417	-4.2	172	0.00
16	Acenaphthene	0.400	0.407	-1.7	164	-0.02
17	Fluorene	0.400	0.403	-0.8	165	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	134	0.00
19	Hexachlorobenzene	0.400	0.395	1.3	146	0.00
20	Pentachlorophenol	0.400	0.862	-115.5#	430	-0.02
21	Phenanthrene	0.400	0.459	-14.8	170	0.00
22	Anthracene	0.400	0.471	-17.7	191	0.00
23	Fluoranthene	0.400	0.621	-55.2#	236	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	123	-0.02
25	Pyrene	0.400	0.792	-98.0#	255	-0.02
26 S	Terphenyl-d14	0.400	0.533	-33.3#	174	0.00
27	Benzo(a)anthracene	0.400	0.920	-130.0#	288	-0.02
28	Chrysene	0.400	0.763	-90.8#	262	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	-5.000	1350.0#	1525	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.422	-5.5	132	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	113	0.00
32	Benzo(b)fluoranthene	0.400	0.753	-88.3#	240	0.00
33	Benzo(k)fluoranthene	0.400	0.593	-48.2#	174	0.00
34 C	Benzo(a)pyrene	0.400	0.564	-41.0#	179	0.00
35	Dibenzo(a,h)anthracene	0.400	0.420	-5.0	134	0.00
36	Benzo(g,h,i)perylene	0.400	0.412	-3.0	125	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 1