

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE091218\
 Data File : BE097461.D
 Acq On : 12 Sep 2018 21:30
 Operator : SJ/JU
 Sample : J4792-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 A3YY2

Manual Integrations
 APPROVED

Sohil
 9/13/2018 5:42:27 PM

Quant Time: Sep 13 02:41:03 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE082918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 13 01:51:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	850	0.40	ng/ul	0.00
2) Naphthalene-d8	10.12	136	4740	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.01	164	3405	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.76	188	8955m	0.40	ng/ul	0.00
16) Chrysene-d12	20.97	240	9881m	0.40	ng/ul	0.00
20) Perylene-d12	23.20	264	9444m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.71	152	1502	0.20	ng/ul	0.00
14) Fluoranthene-d10	18.80	212	5515m	0.19	ng/ul	0.00
Target Compounds					Qvalue	
3) Naphthalene	10.17	128	13354	1.155	ng/ul#	91
5) 2-Methylnaphthalene	11.79	142	4558	0.624	ng/ul	97
7) Acenaphthylene	13.72	152	8042	0.648	ng/ul	98
8) Acenaphthene	14.07	153	4926	0.462	ng/ul	95
9) Fluorene	15.07	166	12980	1.062	ng/ul	92
12) Phenanthrene	16.80	178	110031m	4.773	ng/ul	
13) Anthracene	16.89	178	26818m	1.384	ng/ul	
15) Fluoranthene	18.83	202	107368m	3.626	ng/ul	
17) Pyrene	19.20	202	100125	3.766	ng/ul#	80
18) Benzo(a)anthracene	20.95	228	48936m	1.928	ng/ul	
19) Chrysene	21.00	228	42074	1.439	ng/ul	97
21) Benzo(b)fluoranthene	22.53	252	52757m	2.086	ng/ul	
22) Benzo(k)fluoranthene	22.57	252	14273m	0.506	ng/ul	
23) Benzo(a)pyrene	23.10	252	38259m	1.635	ng/ul	
24) Indeno(1,2,3-cd)pyrene	25.49	276	22363m	0.737	ng/ul	
25) Dibenzo(a,h)anthracene	25.50	278	6067m	0.236	ng/ul	
26) Benzo(g,h,i)perylene	26.19	276	19102m	0.739	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE091218\
Data File : BE097461.D
Acq On : 12 Sep 2018 21:30
Operator : SJ/JU
Sample : J4792-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
A3YY2

Manual Integrations
APPROVED

Sohil
9/13/2018 5:42:27 PM

Quant Time: Sep 13 02:41:03 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE082918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 13 01:51:24 2018
Response via : Initial Calibration

