

Data Path : Z:\HPCHEM1\BNA E\DATA\BE091316\
 Data File : BE092380.D
 Acq On : 13 Sep 2016 12:35
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 13 17:39:52 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	-0.02
2	1,4-Dioxane	0.707	0.730	-3.3	115	-0.01
3	n-Nitrosodimethylamine	0.661	0.644	2.6	133	-0.02
4 S	2-Fluorophenol	1.189	1.193	-0.3	135	0.00
5 S	Phenol-d6	1.586	1.620	-2.1	152#	0.00
6	bis(2-Chloroethyl)ether	1.288	1.210	6.1	136	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
8 S	Nitrobenzene-d5	0.359	0.364	-1.4	155#	0.00
9	Naphthalene	1.120	1.130	-0.9	128	0.00
10	Hexachlorobutadiene	0.175	0.176	-0.6	124	0.00
11	2-Methylnaphthalene	0.733	0.762	-4.0	135	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.02
13 S	2,4,6-Tribromophenol	0.155	0.185	-19.4	185#	0.01
14 S	2-Fluorobiphenyl	1.534	1.477	3.7	133	0.00
15	Acenaphthylene	8.606	8.779	-2.0	142	0.02
16	Acenaphthene	1.393	1.410	-1.2	137	0.00
17	Fluorene	1.749	1.786	-2.1	142	0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	135	0.02
19	Hexachlorobenzene	0.239	0.187	21.8	116	0.00
20	Pentachlorophenol	0.053	0.067	-26.4#	253#	0.00
21	Phenanthrene	1.405	1.141	18.8	121	0.02
22	Anthracene	1.250	1.118	10.6	146	0.02
23	Fluoranthene	6.135	5.482	10.6	136	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
25	Pyrene	4.864	4.871	-0.1	136	0.02
26 S	Terphenyl-d14	0.740	0.709	4.2	131	0.02
27	Benzo(a)anthracene	5.396	5.823	-7.9	142	0.00
28	Chrysene	5.570	5.447	2.2	140	0.00
29	Bis(2-ethylhexyl)phthalate	0.749	0.920	-22.8	192#	0.00
30	Indeno(1,2,3-cd)pyrene	5.684	5.522	2.9	128	0.04
31 I	Perylene-d12	1.000	1.000	0.0	111	0.04
32	Benzo(b)fluoranthene	1.284	1.485	-15.7	145	0.02
33	Benzo(k)fluoranthene	1.357	1.557	-14.7	132	0.04
34 C	Benzo(a)pyrene	1.265	1.391	-10.0	136	0.02
35	Dibenzo(a,h)anthracene	1.170	1.226	-4.8	131	0.05
36	Benzo(g,h,i)perylene	4.933	5.172	-4.8	125	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0