

Data Path : Z:\HPCHEM1\BNA E\DATA\BE091316\
 Data File : BE092380.D
 Acq On : 13 Sep 2016 12:35
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 13 17:39:52 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	115	-0.02
2	1,4-Dioxane	0.400	0.390	2.5	115	-0.01
3	n-Nitrosodimethylamine	0.400	0.396	1.0	133	-0.02
4 S	2-Fluorophenol	0.400	0.402	-0.5	135	0.00
5 S	Phenol-d6	0.400	0.408	-2.0	152	0.00
6	bis(2-Chloroethyl)ether	0.400	0.385	3.8	136	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	121	0.00
8 S	Nitrobenzene-d5	0.400	0.405	-1.3	155	0.00
9	Naphthalene	0.400	0.404	-1.0	128	0.00
10	Hexachlorobutadiene	0.400	0.404	-1.0	124	0.00
11	2-Methylnaphthalene	0.400	0.416	-4.0	135	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	125	0.02
13 S	2,4,6-Tribromophenol	0.400	0.477	-19.2	185	0.01
14 S	2-Fluorobiphenyl	0.400	0.385	3.8	133	0.00
15	Acenaphthylene	0.400	0.408	-2.0	142	0.02
16	Acenaphthene	0.400	0.405	-1.3	137	0.00
17	Fluorene	0.400	0.409	-2.2	142	0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	135	0.02
19	Hexachlorobenzene	0.400	0.312	22.0	116	0.00
20	Pentachlorophenol	0.400	0.559	-39.8#	253	0.00
21	Phenanthrene	0.400	0.325	18.8	121	0.02
22	Anthracene	0.400	0.358	10.5	146	0.02
23	Fluoranthene	0.400	0.357	10.8	136	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	129	0.00
25	Pyrene	0.400	0.401	-0.3	136	0.02
26 S	Terphenyl-d14	0.400	0.384	4.0	131	0.02
27	Benzo(a)anthracene	0.400	0.432	-8.0	142	0.00
28	Chrysene	0.400	0.330	17.5	140	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.457	-14.2	192	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.389	2.8	128	0.04
31 I	Perylene-d12	5.000	5.000	0.0	111	0.04
32	Benzo(b)fluoranthene	0.400	0.463	-15.8	145	0.02
33	Benzo(k)fluoranthene	0.400	0.459	-14.8	132	0.04
34 C	Benzo(a)pyrene	0.400	0.440	-10.0	136	0.02
35	Dibenzo(a,h)anthracene	0.400	0.419	-4.7	131	0.05
36	Benzo(g,h,i)perylene	0.400	0.419	-4.7	125	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0