

Data Path : Z:\HPCHEM1\BNA E\DATA\BE091316\
 Data File : BE092382.D
 Acq On : 13 Sep 2016 17:07
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 13 17:43:41 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	-0.02
2	1,4-Dioxane	0.707	0.797	-12.7	119	-0.01
3	n-Nitrosodimethylamine	0.661	0.666	-0.8	131	-0.02
4 S	2-Fluorophenol	1.189	1.133	4.7	122	0.00
5 S	Phenol-d6	1.586	1.423	10.3	127	0.00
6	bis(2-Chloroethyl)ether	1.288	1.072	16.8	115	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.359	0.351	2.2	132	0.00
9	Naphthalene	1.120	1.134	-1.2	113	0.00
10	Hexachlorobutadiene	0.175	0.183	-4.6	113	0.00
11	2-Methylnaphthalene	0.733	0.724	1.2	113	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.02
13 S	2,4,6-Tribromophenol	0.155	0.167	-7.7	140	0.01
14 S	2-Fluorobiphenyl	1.534	1.517	1.1	114	0.01
15	Acenaphthylene	8.606	8.453	1.8	114	0.02
16	Acenaphthene	1.393	1.426	-2.4	116	0.00
17	Fluorene	1.749	1.761	-0.7	116	0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.02
19	Hexachlorobenzene	0.239	0.188	21.3	98	0.02
20	Pentachlorophenol	0.053	0.057	-7.5	181#	0.00
21	Phenanthrene	1.405	1.148	18.3	103	0.02
22	Anthracene	1.250	1.073	14.2	118	0.02
23	Fluoranthene	6.135	5.517	10.1	116	0.02
24 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
25	Pyrene	4.864	5.115	-5.2	118	0.02
26 S	Terphenyl-d14	0.740	0.754	-1.9	116	0.02
27	Benzo(a)anthracene	5.396	6.575	-21.8	133	0.00
28	Chrysene	5.570	4.686	15.9	100	0.02
29	Bis(2-ethylhexyl)phthalate	0.749	0.762	-1.7	132	0.02
30	Indeno(1,2,3-cd)pyrene	5.684	6.153	-8.3	118	0.05
31 I	Perylene-d12	1.000	1.000	0.0	107	0.03
32	Benzo(b)fluoranthene	1.284	1.306	-1.7	123	0.03
33	Benzo(k)fluoranthene	1.357	1.417	-4.4	116	0.03
34 C	Benzo(a)pyrene	1.265	1.281	-1.3	122	0.03
35	Dibenzo(a,h)anthracene	1.170	1.183	-1.1	122	0.07
36	Benzo(g,h,i)perylene	4.933	5.114	-3.7	120	0.07

(#) = Out of Range

SPCC's out = 0 CCC's out = 0