

Data Path : Z:\HPCHEM1\BNA E\DATA\BE091316\
 Data File : BE092382.D
 Acq On : 13 Sep 2016 17:07
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 13 17:43:41 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	109	-0.02
2	1,4-Dioxane	0.400	0.434	-8.5	119	-0.01
3	n-Nitrosodimethylamine	0.400	0.407	-1.7	131	-0.02
4 S	2-Fluorophenol	0.400	0.381	4.8	122	0.00
5 S	Phenol-d6	0.400	0.359	10.3	127	0.00
6	bis(2-Chloroethyl)ether	0.400	0.350	12.5	115	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.400	0.391	2.3	132	0.00
9	Naphthalene	0.400	0.405	-1.3	113	0.00
10	Hexachlorobutadiene	0.400	0.418	-4.5	113	0.00
11	2-Methylnaphthalene	0.400	0.395	1.3	113	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	104	0.02
13 S	2,4,6-Tribromophenol	0.400	0.432	-8.0	140	0.01
14 S	2-Fluorobiphenyl	0.400	0.395	1.3	114	0.01
15	Acenaphthylene	0.400	0.393	1.8	114	0.02
16	Acenaphthene	0.400	0.409	-2.2	116	0.00
17	Fluorene	0.400	0.403	-0.8	116	0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	114	0.02
19	Hexachlorobenzene	0.400	0.314	21.5	98	0.02
20	Pentachlorophenol	0.400	0.489	-22.2	181	0.00
21	Phenanthrene	0.400	0.327	18.3	103	0.02
22	Anthracene	0.400	0.343	14.2	118	0.02
23	Fluoranthene	0.400	0.360	10.0	116	0.02
24 I	Chrysene-d12	5.000	5.000	0.0	107	0.00
25	Pyrene	0.400	0.421	-5.2	118	0.02
26 S	Terphenyl-d14	0.400	0.408	-2.0	116	0.02
27	Benzo(a)anthracene	0.400	0.487	-21.7	133	0.00
28	Chrysene	0.400	0.268	33.0#	100	0.02
29	Bis(2-ethylhexyl)phthalate	0.400	0.392	2.0	132	0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.433	-8.2	118	0.05
31 I	Perylene-d12	5.000	5.000	0.0	107	0.03
32	Benzo(b)fluoranthene	0.400	0.407	-1.7	123	0.03
33	Benzo(k)fluoranthene	0.400	0.418	-4.5	116	0.03
34 C	Benzo(a)pyrene	0.400	0.405	-1.3	122	0.03
35	Dibenzo(a,h)anthracene	0.400	0.404	-1.0	122	0.07
36	Benzo(g,h,i)perylene	0.400	0.415	-3.7	120	0.07

(#) = Out of Range

SPCC's out = 0 CCC's out = 0