

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091415\
 Data File : BE090796.D
 Acq On : 14 Sep 2015 10:37
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 15 07:53:12 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	109	0.00
2	1,4-Dioxane	1.000	0.846	15.4	89	0.00
3	n-Nitrosodimethylamine	1.000	0.857	14.3	92	0.00
4 S	2-Fluorophenol	1.000	0.875	12.5	94	0.00
5 S	Phenol-d6	1.000	0.907	9.3	104	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	122	0.00
7 S	Nitrobenzene-d5	1.000	0.856	14.4	105	0.00
8	Nitrobenzene	1.000	0.854	14.6	108	0.00
9	Naphthalene	1.000	0.876	12.4	108	0.00
10	Hexachlorobutadiene	1.000	0.849	15.1	100	0.00
11	2-Methylnaphthalene	1.000	0.938	6.2	119	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	138	0.00
13 S	2,4,6-Tribromophenol	1.000	0.870	13.0	123	0.00
14 S	2-Fluorobiphenyl	1.000	0.835	16.5	117	0.00
15	Acenaphthylene	1.000	0.859	14.1	124	0.00
16	Acenaphthene	1.000	0.872	12.8	119	0.00
17	Fluorene	1.000	0.873	12.7	122	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	135	0.00
19	4-Bromophenyl-phenylether	1.000	0.858	14.2	121	-0.02
20	Hexachlorobenzene	1.000	0.829	17.1	112	0.00
21	Pentachlorophenol	1.000	0.964	3.6	106	0.00
22	Phenanthrene	1.000	0.907	9.3	123	-0.02
23	Anthracene	1.000	0.917	8.3	129	0.00
24	Fluoranthene	1.000	0.931	6.9	136	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	143	0.00
26	Pyrene	1.000	1.032	-3.2	155	0.00
27 S	Terphenyl-d14	1.000	0.960	4.0	143	0.00
28	Benzo(a)anthracene	1.000	0.954	4.6	138	0.00
29	Chrysene	1.000	0.949	5.1	131	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.993	0.7	159	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.986	1.4	152	-0.01
32 I	Perylene-d12	5.000	5.000	0.0	155	0.00
33	Benzo(b)fluoranthene	1.000	0.890	11.0	142	0.00
34	Benzo(k)fluoranthene	1.000	0.918	8.2	144	0.00
35 C	Benzo(a)pyrene	1.000	0.939	6.1	156	0.00
36	Dibenzo(a,h)anthracene	1.000	0.921	7.9	150	-0.01
37	Benzo(g,h,i)perylene	1.000	0.876	12.4	143	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0