

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091415\
 Data File : BE090796.D
 Acq On : 14 Sep 2015 10:37
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 15 07:53:12 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.814	0.647	20.5	89	0.00
3	n-Nitrosodimethylamine	1.033	0.819	20.7	92	0.00
4 S	2-Fluorophenol	1.039	0.837	19.4	94	0.00
5 S	Phenol-d6	1.392	1.263	9.3	104	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
7 S	Nitrobenzene-d5	0.330	0.283	14.2	105	0.00
8	Nitrobenzene	0.365	0.309	15.3	108	0.00
9	Naphthalene	1.091	0.955	12.5	108	0.00
10	Hexachlorobutadiene	0.171	0.145	15.2	100	0.00
11	2-Methylnaphthalene	0.590	0.553	6.3	119	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	138	0.00
13 S	2,4,6-Tribromophenol	0.143	0.107	25.2#	123	0.00
14 S	2-Fluorobiphenyl	1.385	1.156	16.5	117	0.00
15	Acenaphthylene	7.992	6.863	14.1	124	0.00
16	Acenaphthene	1.335	1.165	12.7	119	0.00
17	Fluorene	1.666	1.454	12.7	122	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	135	0.00
19	4-Bromophenyl-phenylether	0.170	0.146	14.1	121	-0.02
20	Hexachlorobenzene	0.928	0.759	18.2	112	0.00
21	Pentachlorophenol	0.054	0.037	31.5#	106	0.00
22	Phenanthrene	1.237	1.122	9.3	123	-0.02
23	Anthracene	1.035	0.949	8.3	129	0.00
24	Fluoranthene	1.304	1.215	6.8	136	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	143	0.00
26	Pyrene	1.035	1.068	-3.2	155#	0.00
27 S	Terphenyl-d14	0.553	0.531	4.0	143	0.00
28	Benzo(a)anthracene	1.156	1.059	8.4	138	0.00
29	Chrysene	1.279	1.165	8.9	131	0.00
30	Bis(2-ethylhexyl)phthalate	0.345	0.283	18.0	159#	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	1.164	1.4	152#	-0.01
32 I	Perylene-d12	1.000	1.000	0.0	155#	0.00
33	Benzo(b)fluoranthene	1.084	0.965	11.0	142	0.00
34	Benzo(k)fluoranthene	1.263	1.159	8.2	144	0.00
35 C	Benzo(a)pyrene	0.970	0.911	6.1	156#	0.00
36	Dibenzo(a,h)anthracene	1.020	0.951	6.8	150#	-0.01
37	Benzo(g,h,i)perylene	1.116	0.977	12.5	143	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0